

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2022-2023/100689

03rd Sep 2022

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H500** bearing serial number: **902YOXH02113** installed at **DR. MANISHA MISHRA'S PATHOLOGY LABS, GONDIA** as calibrated on 03rd Sep 2022

Calibrator : ABX MINOCAL

Lot No. : CX474

Expiry Date : 05th Oct 2022

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 02nd Sep 2023.



Shrish Dixit

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**