

JERATH PATH LAB & ALLERGY TESTING CENTRE (757)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 02

Report Date : 31/08/2022

Instrument : Meril Diagnostics AutoQuant 200i

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	82.56	--	1.58	1.50
@ Creatinine	mg/dl	0.92	--	0.71	0.48
@ Bilirubin Total	mg/dl	0.60	--	-1.40	-1.52
@ Bilirubin Direct	mg/dl	0.06	--	-1.71	-1.35
@ Total Protein	g/dL	6.15	--	-0.20	0.88
@ Albumin	g/dL	3.99	--	-0.24	-0.23
@ Calcium	mg/dl	7.95	--	1.38	1.94
@ Triglycerides	mg/dl	131.23	--	1.42	2.09
! ALT	U/L	11.88	--	2.23	0.87
@ AST	U/L	17.05	--	-1.50	-0.75
@ ALP	U/L	42.37	--	-0.75	-0.67

Legend @ : z score  $\leq$  2.0 - Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score  $\geq$  3 - Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	11
Not Evaluated Parameters	0
Evaluated Parameters	11
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A,  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

End of Report