

JERATH PATH LAB & ALLERGY TESTING CENTRE (757)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 02

Report Date : 31/08/2022

Instrument : Sysmex XP-100

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	5.07	5.38	-1.11	-0.91
@ Total Leucocytes (WBC) count	/c.mm	3900.00	4,300.00	-0.29	-0.14
@ Haemoglobin	g/dL	16.60	15.98	0.79	-0.52
@ MCV (Mean Corpuscular Volume)	fL	93.10	93.48	-0.19	-0.34
! MCH (Mean Corpuscular Hb)	pg	32.70	29.72	2.11	0.92
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	35.20	32.40	1.20	0.64
@ Platelet Count	1000/ μ L	152.00	151.23	0.15	0.36
@ Hematocrit	%	47.20	49.29	-0.63	-0.75
@ RDW CV%	%	14.50	14.80	-0.24	-0.26

Legend @ : z score \leq 2.0 - Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score \geq 3 - Unacceptable
 # : Not Evaluated
 Ⓞ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416,
 Commercial Building - 1A,
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

End of Report