

## CALIBRATION CERTIFICATE

To,

This is to Certify that Biochemistry Analyzer Meril **CliniQuant Micro** Instrument Sr.No. **CM0921150** Installed at **Renuka Diagnostics Pathology - Bilaspur** Add: 353, Near Maharana Pratap Chowk, Ring Road No. 2, Gauvrav Path, Bilaspur, Chattisgarh. 495001 \_ has been calibrated by our engineer on **02-Aug -2022** with appropriate and sophisticated tools. Bellow parameters are been calibrated:

### TEST ITEMS AND RESULT

Appearance	OK
AD Value Range (120<AD<350)	OK
Temperature Checking (at 37'C)	OK
Printer Test	OK
Aspiration Volume Range	OK
Stability of Pot. Dichromate	OK
Absorbency Linearity at 340	OK
PC Communication Checking	OK
Sample Test 1	OK
Sample Test 2	OK
Sample Test 3	OK

This Calibration Certificate is Valid Up to **01-Aug-2023** .

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

**Engineer Name - Mr. Mayank Vaishnav.**

**Designation -Service Engineer.**

Thanks and Regards,



For Meril Diagnostics Pvt. Ltd.

**Meril Diagnostics Pvt. Ltd.**

**Corporate Office:** Survey No. 135/ 139, Bilakhia House, Muktanand Marg, Chala, **Vapi** – 396191, **Gujarat**

**Mumbai Office:** 602, Midas, Sahar Plaza, JB Nagar, Andheri East, **Mumbai** – 400059, **Maharashtra**

**Call Centre: 18002663745 Direct Line: +91-22-30732544 E mail: [diagnostics.csd@merillife.in](mailto:diagnostics.csd@merillife.in)**

T +91 260 3052100  
F +91 260 3052125

[www.merillife.com](http://www.merillife.com)

Meril Diagnostics Pvt. Ltd.  
Survey No.135/2/B, Bilakhia House  
Muktanand Marg, Chala  
Vapi - 396191, Gujarat, India.