



Lab 131480
VEDANTA HOSPITAL
BILARIAGANJ ROAD
AZAMGARH, UTTAR
PARDESH, 276001,
INDIA

Sample Summary Report
Hematology Program

Cycle 16
Sep 2022 - Aug 2022
Sample No: 8
Sample Date: 08 Mar 21
Lot No: 261500



Instrument: Sysmex XN-L Series

Analyte	Unit	Result	Mean	Z-score	RMZ	Comparator
✓ Hematocrit	%	44.6	45.1	-0.56	-0.52	Peer
✓ Hemoglobin, Total	g/dL	15.4	15.6	-1.26	-1.43	Peer
✓ MCH	pg/cell	30.2	29.8	0.93	-0.59	Peer
✓ MCHC	g/dL	34.5	34.6	-0.10	-0.18	Peer
✓ MCV	fL (cubic μ m)	87.5	86.2	0.95	-0.24	Peer
✓ MPV	fL (cubic μ m)	10.8	10.5	0.88	0.62	Peer
✓ Platelets	K/ μ L	229	229	0.02	0.26	Peer
▽ RDW	%	14.8	14.2	2.02	1.35	Peer
✓ RDW-SD	fL (cubic μ m)	44.8	43.6	0.99	0.28	Peer
✓ Red Blood Cells	M/ μ L	5.1	5.23	-1.79	-0.46	Peer
✓ WBC	K/ μ L	15.61	16.0	-1.13	-0.63	Peer

Legend: ✓ No Warnings △ Missing Result 1 Late Results ▽ $2.0 < |Z\text{-score}| < 3.0$ ✕ $|Z\text{-score}| > 3.0$
 * Amended Result (per participant's request) = Non-robust determination of Mean and SD

Problem Classification:

Corrective Action:

Reviewed by:

Date:

Copyright © 2021 Bio-Rad Laboratories, Inc. All rights reserved.
 EQAS is a trademark of Bio-Rad Laboratories, Inc.

Expanded uncertainty coverage factor $k=2$
 Generated 11 Nov 2021