

### CALIBRATION CERTIFICATE

**NAME OF CUSTOMER** : OLI CLINICAL LABORATORY  
EAST MAIN ROAD,  
PENNADAM-606105

**UNIT DESCRIPTION** : REFRIGERATOR

**MODEL** : BR 1852

**DATE OF CALIBRATION** : 13.09.2022

**VALID UP TO** : 12.09.2023

**INSTRUMENT USED FOR CALIBRATION:** DIGITAL THERMOMETER

**MODEL NO:**KUSAM-MECO&KM-945

**SL.NO:** 0121849

**CERTIFIED BY:** MSIR INDIA PVT. LTD

**INSTRUMENT VALIDITY:** 22-02-2023

**ENVIROMENTAL CONDITION:**

S.NO	Instrument Temperature Displayed	Thermometer Temperature Displayed	Difference $\pm 1^{\circ}\text{C}$	Remarks
1.	TOP - $+4^{\circ}\text{C}$	$+4.1^{\circ}\text{C}$	0.1	OK
2.	MIDDLE - $+4^{\circ}\text{C}$	$+4^{\circ}\text{C}$	0.0	OK
3.	BOTTOM - $+4^{\circ}\text{C}$	$+4.3^{\circ}\text{C}$	0.3	OK

**STANDARD INSTRUMENT USED:**

<b>Nomenclature:</b>	Digital Thermometer with Sensor(K Type)	<b>Serial no:</b>	S1023388
<b>Make&amp; Model:</b>	KUSAM MECO & KM-945	<b>Code/ID No:</b>	---
<b>Range:</b>	$-50^{\circ}\text{C}$ to $1300^{\circ}\text{C}$	<b>Resolution:</b>	$1^{\circ}\text{C}$

**TESTED BY:** G.DINESH

**APPROVED BY:** G DINESH

**NOTE:** 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.

