

CALIBRATION REPORT FOR DIGITAL THERMOMETER

Certificate Issued to : OLI CLINICAL LABORATORY
EAST MAIN ROAD,
PENNADAM-606105

Company : SIGMA
Serial no : 1263
Date of calibration : 13-09-2022
Valid up to : 12-09-2023



ENVIRONMENTAL CONDITION:

S.NO	Instrument Temperature Displayed	Thermometer Temperature Displayed	Difference $\pm 1^{\circ}\text{C}$	Remarks
1.	TOP +3.8 $^{\circ}\text{C}$	+3.6 $^{\circ}\text{C}$	0.2 $^{\circ}\text{C}$	OK
2.	MIDDLE +3.4 $^{\circ}\text{C}$	+3.5 $^{\circ}\text{C}$	0.1 $^{\circ}\text{C}$	OK
3.	BOTTOM+4 $^{\circ}\text{C}$	+3.9 $^{\circ}\text{C}$	0.1 $^{\circ}\text{C}$	OK

STANDARD INSTRUMENT USED:

Nomenclature:	Digital Thermometer with Sensor (KTYPE)	Serial no:	S1023398
Make& Model:	KUSAM MECO & KM-946	Code/ID No:	---
Range:	(-50°C TO 100°C)	Resolution:	1°C

TEST BY: G.DINESH



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.