

**CALIBRATION REPORT FOR DIGITAL THERMOMETER**

Certificate Issued to : OLI CLINICAL LABORATORY  
EAST MAIN ROAD,  
PENNADAM-606105

Company : SIGMA  
Serial no : 0542  
Date of calibration : 13-09-2022  
Valid up to : 12-09-2023



**ENVIROMENTAL CONDITION:**

S.NO	Instrument Temperature Displayed	Thermometer Temperature Displayed	Difference $\pm 1^{\circ}\text{C}$	Remarks
1.	TOP +3.8 $^{\circ}\text{C}$	+3.6 $^{\circ}\text{C}$	0.2 $^{\circ}\text{C}$	OK
2.	MIDDLE +3.4 $^{\circ}\text{C}$	+3.5 $^{\circ}\text{C}$	0.1 $^{\circ}\text{C}$	OK
3.	BOTTOM+4 $^{\circ}\text{C}$	+3.9 $^{\circ}\text{C}$	0.1 $^{\circ}\text{C}$	OK

**STANDARD INSTRUMENT USED:**

<b>Nomenclature:</b>	Digital Thermometer with Sensor (KTYPE)	<b>Serial no:</b>	S1023398
<b>Make&amp; Model:</b>	KUSAM MECO & KM-946	<b>Code/ID No:</b>	---
<b>Range:</b>	(-50°C TO 100°C)	<b>Resolution:</b>	1°C

TEST BY: G.DINESH



NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should be brought to our lab within 45 days from the date of issue of certificate.