

CALIBRATION CERTIFICATE

CUSTOMER ADDRESS:	
M/S.	
OLI CLINICAL LABORATORY	
EAST MAIN ROAD,	
PENNADAM-606105	
	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT: MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : NBS210847251	DATE OF ISSUE : 13.09.2022
RECEIPT NO : 10136	TEMPERATURE : 24.5°C
RECEIPT DATE : 13.09.2022	HUMIDITY : 53%
JOB CARD NO : TPM000	AIR PRESURE : 959hpa
JOB CARD DATE : 13.09.2022	MEDIUM OF WEIGHING : GRADE III WATER
	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 10-100 µl	DENSITY CORRECTION FACTOR[Z]: : 1.0037 µl/mg
TIPS USED : 10-100µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE ACCREDITATION: VOLUME-MICROPIPETTE (1µl-10000µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT	: SHIMADZU
SL.NO	: D450020324
RESOLUTION	: 0.00001g upto 82g &above is 0.0001 g
CALIBRATED BY	: NEWTONBIOMEDICAL SERVICE
CALIBRATION VALIDITY	: 12.09.2023

