

NEWTON



BIOMEDICAL SERVICE

GSTIN : 33AAOFN5502A1ZR

CALIBRATION CERTIFICATE

| | |
|--|---------------------------------------|
| CUSTOMER ADDRESS: | |
| M/S. | |
| OLI CLINICAL LABORATORY | |
| EAST MAIN ROAD, | |
| PENNADAM-606105 | |
| | DISCIPLINE : MECHANICAL CALIBRATION |
| NAME OF EQUIPMENT: MICROPIPETTE | PARAMETER : VOLUME |
| CERTIFICATE NO : NBS210847251 | DATE OF ISSUE : 25-03.2021 |
| RECEIPT NO : 10136 | TEMPERATURE : 24.5 ⁰ C |
| RECEIPT DATE : 13.09.2022 | HUMIDITY : 53% |
| JOB CARD NO : TPM000 | AIR PRESURE : 959hpa |
| JOB CARD DATE : 25-03-2021 | MEDIUM OF WEIGHING : GRADE III WATER |
| | DENSITY OF MEDIUM : APPROXIMATELY 1 |
| RANGE : 20 μ l | DENSITY CORRECTION |
| | FACTOR[Z]: : 1.0037 μ l/mg |
| | |
| TIPS USED : 10-100 μ l | WEIGHING BALANCE : SEMI MICRO BALANCE |
| SCOPE ACCREDITATION: VOLUME-MICROPIPETTE (1 μ l-10000 μ l) | |

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

| | |
|----------------------|---|
| MASTER EQUIPMENT | : SHIMADZU |
| SL.NO | : D450020324 |
| RESOLUTION | : 0.00001g upto 82g & above is 0.0001 g |
| CALIBRATED BY | : NEWTONBIOMEDICAL SERVICE |
| CALIBRATION VALIDITY | : 12.09.2023 |

