



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3272

Distribution No.: 156-I

Month/Year: June/2022

Instrument ID: PENTRA E S 60 SERIAL NO. 603PES1405Q

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 29-08-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.7	5.6	11.3	10.88	0.1010	0.25	0.1	0.1	0.0120	0.00
RBC x10 <sup>6</sup> /µl	1	4.55	4.4	8.95	8.86	0.0140	0.38	0.15	0.04	0.0040	2.97
Hb g/dl	1	13.1	13	26.1	25.4	0.0420	1.05	0.1	0.1	0.0110	0.00
HCT%	1	39.7	38.9	78.6	80.35	0.2230	-0.44	0.8	0.4	0.0360	1.14
MCV-fl	1	88	87	175	182.5	0.3910	-1.07	1	0.3	0.0370	1.40
MCH-Pg	1	29.6	28.9	58.5	57.35	0.1040	0.71	0.7	0.2	0.0230	2.25
MCHC-g/dl	1	33.5	33.1	66.6	62.9	0.1840	0.99	0.4	0.3	0.0220	0.45
Plt. x10 <sup>3</sup> /µl	1	249	245	494	418.5	2.67	1.59	4	5	0.52	-0.17
Retic %	2	8	7	15	10.4	0.40	0.59	1	0.5	0.05	0.84

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
<b>DLC%</b>	3	Nrbcs=1 , Poly=10 L=15, E=0, Mono/Promono=6 , B1=64 P.M.=1, Mye=2, Meta=2, Other=
<b>RBC Morphology</b>	3	Predominantly: Normocytic/ Normochromic, Moderate: Anisocytosis, Microcytic
<b>Diagnosis</b>	3	Acute myeloid leukemia . Advice: immunophenotyping by flow cytometry for further sub-categorization.

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 156--I	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	138	138	87.68	83.33	5.8	2.9	6.52	13.77
<b>RBC x10<sup>6</sup>/µl</b>	1	138	138	87.68	92.75	7.25	4.35	5.07	2.9
<b>Hb g/dl</b>	1	138	138	92.03	88.41	5.8	5.8	2.17	5.79
<b>HCT%</b>	1	138	138	89.86	90.58	7.25	5.8	2.89	3.62
<b>MCV-fl</b>	1	138	138	93.48	94.2	4.35	1.45	2.17	4.35
<b>MCH-Pg</b>	1	138	138	86.23	90.58	10.87	5.8	2.9	3.62
<b>MCHC-g/dl</b>	1	138	138	93.48	90.58	5.07	2.9	1.45	6.52
<b>Plt. x10<sup>3</sup>/µl</b>	1	138	138	93.48	92.03	4.35	4.35	2.17	3.62
<b>ReticCount%</b>	2	138	115	96.52	89.57	0.00	6.96	3.48	3.47
<b>PS Assessment</b>	3	138	126	Satisfactory :91.31%, Borderline Sat. :0.72%, Unsatisfactory :7.97%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----