



# APEX QUALITY CERTIFICATION SERVICES PVT. LTD.



Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019  
Ph.: 0141-4036776, 9414079483, 9636293333  
Email: apex.qcs@gmail.com

## CALIBRATION CERTIFICATE

Page 01 of 01

CERTIFICATE NO.: AQCS/22091301.10001		Date of Calibration : 13/09/2022	
		Recommended due date : 12/09/2023	
<b>CUSTOMER'S DETAILS</b>		<b>INSTRUMENT DESCRIPTION</b>	
M/S Shree Rugha Ram Bhadu Memorial Education And Welfare Society Vill.: Seruna, Tehsil: Shree Dungargarh, Dist.: Bikaner		Nomenclature : Centrifuge Machine Make / Model : Remi /R-303 Sr. No./Ident. No. : ZHHN-28454 Range : 0-3800 RPM Least Count : NA Location : -- Cal. Performed At : Lab	
Service Request No. : 220711/01 Dated: 11/07/2022			
<b>STANDARD EQUIPMENT USED:</b>			
Nomenclature	Identification No.	Traceability Certificate No.	Valid up to
Digital Tachometer	S999205	22000022012	08/06/2023
<b>ENVIRONMENTAL CONDITIONS:</b>		<b>PROCEDURE FOLLOWED:</b>	
Temperature : 23 ± 5°C Relative Humidity : 50± 10%		CP/RM-01	
<b>CALIBRATION RESULT</b>			
S. No.	Set Knob on DUC	Measured Value on Standard RPM	
1	1	1115	
2	2	1918	
3	3	2696	
4	4	3175	
5	5	3795	
<b>REMARKS :</b>			
1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is ± 2.37% of reading.			
2. The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.			
3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur			
4. Format Number.: F-06/00			
Calibrated by  (Sachin Yadav)		Approved by  ( R. R. Yadav ) Quality Manager	

