



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4484

Distribution No.: 155-L

Month/Year: April/2022

Instrument ID: HUMACOUNT 5D

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 04-07-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	14.7	14.2	28.9	24.12	0.1270	1.44	0.5	0.2	0.0130	1.32
RBC x10 ⁶ /µl	1	6.13	6.12	12.25	12.05	0.0170	0.44	0.01	0.06	0.0040	-0.84
Hb g/dl	1	14.3	14.3	28.6	28.5	0.0290	0.12	0	0.1	0.0080	-1.35
HCT%	1	48.9	48.9	97.8	95.8	0.2590	0.26	0	0.5	0.0360	-0.84
MCV-fl	1	79.9	79.77	159.67	159.6	0.3180	0.01	0.13	0.3	0.0210	-0.46
MCH-Pg	1	23.37	23.33	46.7	47.1	0.0570	-0.26	0.04	0.2	0.0120	-1.08
MCHC-g/dl	1	29.24	29.24	58.48	58.9	0.1470	-0.10	0	0.25	0.0170	-0.84
Plt. x10 ³ /µl	1	195	184	379	399	2.71	-0.25	11	9	0.55	0.21
Retic %	2										

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=02 , Poly=71 L=04, E=00, Mono/Promono=01 , B1=00 P.M.=00, Mye=16, Meta=08, Other=00
RBC Morphology	3	Poly: 50 - 66, Myelo: 9 - 18, Meta: 6 - 13, Lympho: 3-7, nRBC/Promyelo/Blast/Eos/Baso/Mono: 0 - 5
Diagnosis	3	CHRONIC MYELOPROLIFERATIVE NEOPLASM ? CML
		Chronic Myeloid Leukemia (Chronic Phase)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 155--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	347	334	87.43	89.22	5.69	4.19	6.88	6.59
RBC x10⁶/µl	1	347	347	83.29	86.46	7.49	2.02	9.22	11.52
Hb g/dl	1	347	347	87.32	84.73	4.9	6.05	7.78	9.22
HCT%	1	347	333	90.39	88.89	6.61	5.11	3	6
MCV-fl	1	347	333	90.09	86.19	5.41	3.9	4.5	9.91
MCH-Pg	1	347	333	87.99	93.69	7.51	3	4.5	3.31
MCHC-g/dl	1	347	333	91.59	86.79	4.8	4.5	3.61	8.71
Plt. x10³/µl	1	347	333	96.7	86.19	1.8	8.71	1.5	5.1
ReticCount%	2	347	222	91.44	93.24	5.41	2.7	3.15	4.06
PS Assessment	3	347	230	Satisfactory :96.26%, Borderline Sat. :2.88%, Unsatisfactory :0.86%					

***Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----