



CALIBRATED ITEM

Multitech Technofab Enterprises

Item's Name: **CENTRIFUGE**

Cal. on: **03/06/2022** Due on: **03/06/2023**

Sr. / I. D. No.: **OHC/INST/CEN/03**

Job I. D. No.: **MTE/EN-1208/8-255**

Ph.: **011-41764006, 09810943334**

Enterprises

ificate No.: **CC-3225**

ite, Patparganj, Delhi -110092

3334, 8920678805

delhi2020@gmail.com

fab.com



CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/6594 URL : CC32252200001932F Calibration Issued on: 04-06-2022

Job / Identification No. : MTE/CAL/JOB/8255 Date of Calibration : 03-06-2022

Job Request Date: 31-05-2022 Recommended Date of Next Calibration: 03-06-2023

Calibrated For: Orbit Health Care and Diagnostic Centre
325/11 Near Aryan Hospital, Old Railway Road, Gurugram Haryana Reference : Order thru Telephonic

Discipline: Mechanical Group: Acceleration & Speed

Certificate for : Centrifuge Item Location : LAB

Make / Model Serial / I.D. No. Inst. Range Unit Item Condition

LABRON / NA NA / OHC/INST/CEN/03 0 to 3500 RPM Working

Environmental Condition 25.8 ± 0.5 °C 51 ± 3% RH

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard Make / Model Certificate No. Calibration Agency Valid up to

Digital Tachometer(Non contact) Lutron / DT-2234C AACPL/01264F A A Calibration Pvt. Ltd. 17-01-2023

Calibration Procedure Lab Procedure : MTE-2/RPM/2.50 STD Procedure : NA

RESULTS: On-Site Calibration
All Readings are in RPM, Specified Otherwise

U.U.C. READING(Set At)	Average	UNCERTAINTY^ in (%)
Knob Set at 1	504.7	± 2.31
Knob Set at 2	1308	± 2.31
Knob Set at 3	1998.7	± 2.31
Knob Set at 4	2808.7	± 2.31
Knob Set at 5	3303.3	± 2.31

^The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

***** END OF REPORT *****

Calibrated By Sachin
Deputy Quality Manager

For and on behalf of
Multitech Technofab Enterprises
Delhi-110092

Laboratory Head