



Ref : TBMNZ-1/765/22-23

Dated : 18-06-22

Calibration Certificate

Customer Name : Orbit Health Care & Diagnostic Centre, Gurgaon

Model : Erba Chem 5X

Serial No. : S2105131

Calibration Date : 17-06-22

Calibration Due Date : 16-06-23

This is to certify that we have calibrated the Semi Automated Biochemistry Analyser , Model : Erba Chem 5X bearing Sr. No. S2105131 and on inspection of the instrument it is observed that the results are well within the range and instrument is working fine Instrument is properly calibrated

Thanking you,
For **Transasia Bio-Medicals Ltd.**

Varun
Varun Arora
Sr. Service Engineer

ERBA NORM 8C

LOT - S062243 &

EXP - 11/2024.

01/08/22 19:17:03

000108 AST 47.91 IU/l H
LINEAR REACTION

01/08/22 19:23:06

000109 ALT 49.33 IU/l H
LINEAR REACTION

01/08/22 19:29:47

000110 ALP 98.95 IU/l
LINEAR REACTION

01/08/22 19:33:55

01/08/22 20:32:13

000123 URE 41.21 mg/dl
LINEAR REACTION

01/08/22 20:35:15

000124 CRE 1.30 mg/dl
LINEAR REACTION

01/08/22 20:38:17

000125 ALB 2.41 g/dl L

01/08/22 20:42:30

000126 CHO 139.9 mg/dl

01/08/22 20:47:04

000127 TRIG 70.50 mg/dl

01/08/22 20:48:12

000128 TPRO 6.06 g/dl L

01/08/22 20:50:18

000129 URIC 5.25 mg/dl