

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial
Estate, Phase - III 110
020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2022-2023/100654

10th Aug 2022

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX PENTRA XL80** bearing serial number: **209ESOH04697** installed at **Orbit Health Care & Diagnostics Centre, Gurugram** as calibrated on 10th Aug 2022

Calibrator : ABX MINOCAL
Lot No. : CX473
Expiry Date : 05th Sep 2022

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 09th Aug 2023.



Shrish Dixit

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**