



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



PC-1024

Lab Name **DOCOPD LABS** Lab No **16077**
 Constituent Group **Chemistry I** Date of Result Entered : **20/05/2022**
 PT item **Lyophilized Serum** Date of Report Published : **08/06/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Mindray	203	190.00	6.28	11.94	178.4 mg/dL	-0.97	1.68
2	UREA	Urease UV / GLDH	Mindray	224	62.24	6.26	3.90	59.46 mg/dL	-0.71	0.52
3	CREATININE	Enzymatic Colorimetric	Mindray	63	2.54	6.54	0.17	2.48 mg/dL	-0.36	0.04
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Mindray	184	3.55	8.61	0.31	3.65 mg/dL	0.33	0.05
5	T-PROTEIN	Biuret - Colorimetric	Mindray	216	5.88	6.29	0.37	4.57 g/dL	-3.54	0.05
6	ALBUMIN	BCG - colorimetric	Mindray	219	3.56	7.65	0.27	4.09 g/dL	1.95	0.04
7	CALCIUM	Arsenazo III	Mindray	173	11.23	6.66	0.75	12.3 mg/dL	1.43	0.11
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Mindray	139	4.88	14.50	0.71	6.6 mg/dL	2.43	0.12
9	URIC ACID	Enzymatic / Uricase Colorimetric	Mindray	209	4.85	9.64	0.47	4.46 mg/dL	-0.84	0.06
10	CHOLESTEROL	CHOD-PAP	Mindray	232	120.27	6.75	8.12	124.7 mg/dL	0.55	1.07
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	226	110.19	9.91	10.92	110.4 mg/dL	0.02	1.45
12	HDL	Direct method / Enzymatic colorimetric	Mindray	168	26.18	9.65	2.53	25.5 mg/dL	-0.27	0.39
13	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1024	139.25	3.38	4.71	135.6 mmol/L	-0.78	0.29
14	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1002	4.58	5.70	0.26	4.84 mmol/L	1.00	0.02
15	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	785	102.75	4.07	4.18	96.6 mmol/L	-1.47	0.30
16	AST	UV kinetic(with & without PLP (P-5-P))	Mindray	209	66.65	9.02	6.01	52.5 U/L	-2.35	0.83
17	ALT	UV kinetic(with & without PLP (P-5-P))	Mindray	198	51.77	13.54	7.01	73 U/L	3.03	1.00
18	ALP	PNP AMP kinetic	Mindray	148	108.76	13.09	14.23	115 U/L	0.44	2.34
19	AMYLASE	Enzymatic Colorimetric / G7PNP Blocked	Mindray	31	53.92	12.66	6.83	70.5 U/L	2.43	2.45
20	IRON	Ferrene (No Protein Removal)	Any Analyser (Automation / Semi Automation)	106	93.88	10.06	9.44	91 ug/dL	-0.31	1.83
21	MAGNESIUM	Xylidyl Blue	Mindray	23	2.69	11.72	0.32	2.48 mg/dL	-0.67	0.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DOCOPD LABS
B-31 1ST FLOOR SECTOR 6
NOIDA
UTTAR PRADESH201301

Coordinator Contact Details:
Email:clinq@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
******* End of Report *******