



**TBS INDIA**  
**TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.**

**CERTIFICATE OF PERFORMANCE FOR WEIGHING MACHINE**

ISSUED TO : SMHS  
 CERTIFICATE NO : TBS/CAL/GMC/09/22  
 LOCATION/DEPT. : BIOCHEMISTRY DIAGNOSTIC LAB      DATE : 22.09.22  
 ASSET NO : SR-GMC-00330      MODEL NO : BCM 10  
 MANUFACTURER : WEIGHING MACHINE      SERIAL NO : 2GLS28916

**CALIBRATION STANDARD USED**

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	TBS/TE/0381	SMALL WEIGHTS	WENSAR	F1 CLASS	10637	10.03.22	10.03.24
2	TBS/TE/0185	ELECTRICAL SAFETY ANALYZER	RIGEL MEDICAL	288+	25J-0310	24.03.22	23.03.23

**TEST REPORT**

S. No	Parameters	Actual Weight (gm)	Measured Volume (gm)	Acceptable Range	Remarks
1	WEIGHT	0.0000	0.0000	± 2%	PASS
		1.0000	1.0000	± 2%	PASS
		2.0000	1.9888	± 2%	PASS
		5.0000	5.0009	± 2%	PASS
		100.00	99.998	± 2%	PASS
		200.00	199.89	± 2%	PASS

**Electrical Safety Test Passed as per IEC Standard 60601.1**

Calibration Passed:    YES       NO

Calibration Done Date: **22.09.22**

Calibration Due Date: **22.09.23**

Calibrated By : **MOHAMAD ROUMAN**

Approved By: \_\_\_\_\_



*The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards*

5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,  
 Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055  
 Email: [info@tbs-india.com](mailto:info@tbs-india.com) Website: [www.tbs-india.com](http://www.tbs-india.com)