



TBS INDIA

TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.

CERTIFICATE OF PERFORMANCE FOR CENTRIFUGE

ISSUED TO : «SMHS Hospital»
 CERTIFICATE NO : «TBS/CAL/GMC/22/09»
 LOCATION/DEPT. : «BIOCHEMISTRY DIAGNOSTIC LAB» DATE : «22/09/2022»
 ASSET NO : «SR-GMC-00239» MODEL NO : «R8BCL»
 MANUFACTURER : «REMI» SERIAL NO : «NA»

CALIBRATION STANDARD USED

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	«TBS/TE/0385»	DIGITAL TACHOMETER	«METRAVI»	«NCTM-2000»	NIL	«13.1.22»	«13.1.23»
2	«TBS/TE/0185»	«EST ANALYZER»	RIGEL MEDICAL	«288+»	«25J-0310»	«24.3.22»	«23.3.23»

TEST REPORT


S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	SPEED	RPM	1000	992.7	± 5%	PASS
			2000	1999.5	± 5%	PASS
			3000	2996.2	± 5%	PASS
			4000	3995.9	± 5%	PASS
			5900	5899	± 5%	PASS

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed: YES NO

Calibration Done Date: «22.09.22» Calibration Due Date: «27.09.23»

Calibrated By: «MOHAMAD ROUMAN»

Approved By: _____


The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards

**5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,
 Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055
 Email: info@tbs-india.com Website: www.tbs-india.com**