



**TBS INDIA**

**TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.**



**CERTIFICATE OF PERFORMANCE FOR REFRIGIRATOR**

ISSUED TO : SMHS HOSPITAL  
CERTIFICATE NO : «TBS/CAL/GMC/09/22»  
LOCATION/DEPT. : «BIOCHEMISTRY DIAGNOSTIC LAB»» DATE : «22/09/22»  
ASSET NO : «SR-GMC-01981» MODEL NO : NA  
MANUFACTURER : SAMSUNG SERIAL NO : NA

**CALIBRATION STANDARD USED**

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	TBS/TE/0281	DIGITAL THERMOMETER	WACO	201	130230	21.07.22	20.07.23
2	TBS/TE/0185	ELECTRICAL SAFETY ANALYZER	RIGEL MEDICAL	288+	25J-0310	24.03.22	23.03.23

**TEST REPORT**

S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	TEMPERATURE	°C	-17	-17	± 0.3 °C	PASS
			NIL	NIL	NIL	NIL
			NIL	NIL	NIL	NIL
			NIL	NIL	NIL	NIL

**Electrical Safety Test Passed as per IEC Standard 60601.1**

Calibration Passed: YES  NO

Calibration Done Date: «22.09.22»

Calibration Due Date: «22.09.23»

Calibrated By: «MOHAMAD ROUMAN»

Approved By: \_\_\_\_\_



*The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards*

5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,  
Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055  
Email: [info@tbs-india.com](mailto:info@tbs-india.com) Website: [www.tbs-india.com](http://www.tbs-india.com)