



TBS INDIA



TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.

CERTIFICATE OF PERFORMANCE FOR REFRIGIRATOR

ISSUED TO : «SMHS HOSPITAL»
 CERTIFICATE NO : «TBS/CAL/GMC/09/22»
 LOCATION/DEPT. : BIOCHEMISTRY DIAGNOSTIC LAB DATE : «22/09/22»
 ASSET NO : «SR-GMC-01983» MODEL NO : 375 I
 MANUFACTURER : KELVINATOR SERIAL NO : 2003376

CALIBRATION STANDARD USED

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	TBS/TE/0281	DIGITAL THERMOMETER	WACO	201	130230	21.07.22	20.07.23
2	TBS/TE/0185	ELECTRICAL SAFETY ANALYZER	RIGEL MEDICAL	288+	25J-0310	24.03.22	23.03.23

TEST REPORT

S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	TEMPERATURE	°C	10	9	± 0.3 °C	PASS
			NIL	NIL	NIL	NIL
			NIL	NIL	NIL	NIL
			NIL	NIL	NIL	NIL

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed: YES NO

Calibration Done Date: «22.09.22»

Calibration Due Date: «29.09.23»

Calibrated By: «MOHAMAD ROUMAN»

Approved By: _____



The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards

5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,
Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055
Email: info@tbs-india.com Website: www.tbs-india.com