

SHREE DIAGNOSTIC CENTRE

External Quality Assurance Program Evaluation Form

Discipline : Serology

LAB CODE : 2359

EQAS : RML

Sample No : 06

Sr.no	Month : MAY-2022	Test Parameters	Sample Reconstituted & run	Report Received on	Our Results	Acceptable Results (%)	Results/ Z Score/Comments Corrective Action	LD/QM
1		ANTI HCV	Round No 6	20-10-2022	Non Reactive	84%	Within Consensus	Result Satisfactory
2		Dengue IgG	Round No 6	20-10-2022	Non Reactive	94%	Within Consensus	Result Satisfactory
FORM NO: 14								
FORM NAME : EQAS EVALUATION FORM-SEROLOGY								
ISSUE NO: 01								
ISSUE DATE: 01/07/2020								
AMEN NO-00								
AMEND DATE : 00								
ISSUE BY : QM								
APPROVED BY : LD								

FM/EQAS-EVAL-SERO/14



*N. K. Singh*  
DONE BY

*[Signature]*

QUALITY MANAGER



# RML – Quality Assurance Program (RML – QAP)



Lab Code No. **2359**

## BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 11

ROUND: 6

TOTAL PARTICIPANT: 407

DATE: 20/10/2022

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- ANTI HCV	350	Reactive	Reactive : 295 Non-Reactive : 55	84%	Within Consensus
S2- Dengue IgG	312	Non - Reactive	Reactive : 19 Non-Reactive : 293	94%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

Programme Director

Dr. Bandana Mehrotra

Checked By:

Prepared by: SV

**QUALITY MANAGER**

**\*\*End of Report\*\***



Page 1 of 1

Doc. No.: ASS / FR / 05 / R 01 / Dt.: 05.01.2022



Address: B-171, Nirala Nagar, Lucknow - 226 020, Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence