


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221018AVN-105	Page	: 1 of 1	
CALIBRATED ON.	: 18 / 10 / 2022	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 17 / 04 / 2023	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 221018AVN Dt. 18.10.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000684F Dt. 20.10.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. H. No. 31/82A-9-B Rashmi Nagar Near BHU Hospital Varanasi-221005			
ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: (20 to 200) µl
	L. Count	:	ID No.	: RCL/BIO/PIP-05
	Location	: Clinical Biochemistry	S. No.	: SH88520
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023
RESULTS :	Mechanical Discipline			
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		20	20.20	(-) 0.20
200		100	100.38	(-) 0.38
		200	200.46	(-) 0.46

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

