


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 221018AVN-110	<b>Page</b>	: 1 of 1	
<b>CALIBRATED ON.</b>	: 19 / 10 / 2022	<b>Ref. Std.</b>	: IS:12508:1988	
<b>CALIBRATION DUE.</b>	: 18 / 10 / 2023	<b>C.P. No.</b>	: ACS/CP/10	
<b>SRF No. &amp; DATE</b>	: 221018AVN Dt. 18.10.2022	<b>Phy. Inspection</b>	: O.K.	
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000689F Dt. 20.10.2022	<b>Cal. At</b>	: Site / Lab	
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. H. No. 31/82A-9-B Rashmi Nagar Near BHU Hospital Varanasi-221005			
<b>ITEM CALIBRATED</b>	: Centrifuge			
	<b>Make</b>	: Remi	<b>Range</b>	: (0 to 6000) RPM
	<b>L. Count</b>	:	<b>ID No.</b>	: RCL/BIO/Cent-01
	<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: ZIBN-08848
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C + 3°C } R.H. {40% to 60%}			
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	11/09/2023
<b>RESULTS</b>	: Mechanical Discipline			
<b>* U. C. INSTT. READS (RPM)</b>	<b>Standard</b>	<b>Measured</b>	<b>Error</b>	
		<b>(RPM)</b>	<b>(RPM)</b>	
1000		1001.74	- 1.74	
2000		2002.58	- 2.58	
3000		3003.42	- 3.42	
4000		4004.36	- 4.36	
(average of readings)				

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 3.09$  rpm of reading at coverage Factor  $k=2$ .

\*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

