


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



CC-3427

Calibration Certificate

CERTIFICATE NO.	: 221018A/N-111	Page	: 1 of 1
CALIBRATED ON.	: 19 / 10 / 2022	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 18 / 10 / 2023	C.P. No.	: ACS/CP/10
SRF No. & DATE	: 221018A/N Dt. 18.10.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342722000000690F Dt. 20.10.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. H. No. 31/82A-9-B Rashmi Nagar Near BHU Hospital Varanasi-221005		
ITEM CALIBRATED	: Centrifuge	Make	: Remi
		Range	: (0 to 6000) RPM
		L. Count	: ID No. : RCL/CP/Cent-02
		Location	: Clinical Pathology
		S. No.	: ZIBN-08827
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }		
STANDARD INSTRUMENTS USED :			
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.
DUE DATE			
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1
			11/09/2023
RESULTS	: Mechanical Discipline		
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.82	- 1.82	
2000	2002.64	- 2.64	
3000	3003.56	- 3.56	
4000	4004.44	- 4.44	
(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 3.09 rpm of reading at coverage Factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

