

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Triglycerides	Erba Chem 5/Plus	141.20	mg/dl	2.56
X ALT	Erba Chem 5/Plus	24.20	U/L	14.84
! ALP	Erba Chem 5/Plus	68.00	U/L	2.16
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**RIGHT CARE PATHOLOGY LAB (603)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220207 and Sample No 01**


**Report Date : 31/08/2022**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	67.20	--	-0.60	-0.57
@ Urea	mg/dl	19.26	--	-0.18	0.65
@ Creatinine	mg/dl	0.88	0.90	0.27	1.35
@ Bilirubin Total	mg/dl	0.10	--	-0.42	1.49
@ Bilirubin Direct	mg/dl	0.06	--	0.88	2.08
@ Total Protein	g/dL	6.60	--	0.27	-0.35
@ Albumin	g/dL	3.60	--	0.43	0.07
@ Uric Acid	mg/dl	3.90	4.24	-1.87	0.75
@ Calcium	mg/dl	8.60	8.44	0.83	1.04
@ Cholesterol	mg/dl	138.30	--	0.36	-0.56
! Triglycerides	mg/dl	141.20	--	2.56	0.08
@ HDL	mg/dl	39.20	--	1.39	0.38
X ALT	U/L	24.20	--	14.84	2.58
! ALP	U/L	68.00	100.85	2.16	2.56
@ LDH	U/L	232.40	200.80	1.57	0.02
# Chloride	mmol/L				
@ Sodium	mmol/L	135.00	--	-1.03	-1.47
@ Potassium	mmol/L	5.10	--	0.71	2.18
@ C-Reactive Protein Quantitative	mg/L	7.30	7.26	-0.10	1.15

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	1
Evaluated Parameters	18
Outlier Parameters (X)	1
EQAS Score Biochemistry	94.44 %



**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kiroli Road, Kuria (W),  
Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Bilirubin Direct	Erba Chem 5/Plus	0.70	mg/dl	*
X Albumin	Erba Chem 5/Plus	5.13	g/dL	7.78
! Uric Acid	Erba Chem 5/Plus	6.00	mg/dl	2.28
X ALP	Erba Chem 5/Plus	93.50	U/L	5.33
X C-Reactive Protein Quantitative	Erba Chem 5/Plus	10.04	mg/L	4.07

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**RIGHT CARE PATHOLOGY LAB (603)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220104 and Sample No 01**

**Report Date : 23/05/2022**

**Instrument : Erba Chem 5/Plus**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	82.28	66.54	1.81	-0.49
@ Urea	mg/dl	21.02	22.71	-0.22	1.07
@ Creatinine	mg/dl	0.90	0.89	0.38	1.85
* Bilirubin Direct	mg/dl	0.70		*	1.01
@ Total Protein	g/dL	7.63	7.34	1.83	-0.51
X Albumin	g/dL	5.13		7.78	0.14
! Uric Acid	mg/dl	6.00	4.96	2.28	1.60
@ Calcium	mg/dl	8.07	8.45	0.02	0.96
@ Cholesterol	mg/dl	157.00	153.57	1.68	-0.74
@ Triglycerides	mg/dl	109.00	120.45	0.39	-0.54
@ HDL	mg/dl	32.60	41.05	-0.44	-0.08
@ ALT	U/L	9.78	11.43	0.72	-0.99
X ALP	U/L	93.50	--	5.33	2.80
@ LDH	U/L	235.20	--	-0.64	-0.89
@ Chloride	mmol/L	109.00	--	0.22	-3.11
@ Sodium	mmol/L	145.00	--	1.52	-1.72
@ Potassium	mmol/L	4.35	--	0.50	2.90
X C-Reactive Protein Quantitative	mg/L	10.04	7.10	4.07	1.86

Legend @ : z score  $\leq$  2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score  $\geq$  3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	4
EQAS Score Biochemistry	77.78 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**

Unit No. 409-416,  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070

End of Report