

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Erythrocyte (RBC) Count	AGD300	1.15	mill/cu.mm	-2.85
! Platelet Count	AGD300	27.00	1000/ μ L	-2.61
<p>Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated 🕒 : Delayed Result Entry * : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RIGHT CARE PATHOLOGY LAB (603)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 03


Report Date : 31/08/2022

Instrument : AGD300

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Erythrocyte (RBC) Count	mill/cu.mm	1.15	1.15	-2.85	-1.06
@ Total Leucocytes (WBC) count	/c.mm	1600.00	1,540.00	0.48	-0.42
@ Haemoglobin	g/dL	2.80	2.65	-1.16	-0.88
@ MCV (Mean Corpuscular Volume)	fL	109.60	96.25	-0.43	-0.17
@ MCH (Mean Corpuscular Hb)	pg	24.30	25.15	-0.14	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	22.20	26.80	0.23	0.24
! Platelet Count	1000/ μ L	27.00	27.00	-2.61	-1.04
@ Hematocrit	%	12.60	10.30	-0.87	-0.69
@ RDW CV%	%	17.10	--	-1.18	-1.20

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RIGHT CARE PATHOLOGY LAB (603)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 01


Report Date : 31/08/2022

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	40-70%	40-70%
@ Lymphocyte.	-	20-40%	20-40%
@ Eosinophil.	-	<6%	<6%

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry

Total Parameters	3
Not Evaluated Parameters	0
Evaluated Parameters	3
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroli Road, Kurla (W),
 Mumbai - 400070

End of Report

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Platelet Count	AGD300	198.00	1000/ μ L	-2.27
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

RIGHT CARE PATHOLOGY LAB (603)

Outlier And Analyte Summary Report

Result Details For Cycle No 220207 and Sample No 04


Report Date : 31/08/2022

Instrument : AGD300

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.58	2.63	-1.46	-1.06
@ Total Leucocytes (WBC) count	/c.mm	10300.00	10,145.00	0.47	-0.42
@ Haemoglobin	g/dL	7.80	7.45	-0.82	-0.88
@ MCV (Mean Corpuscular Volume)	fL	127.20	101.80	-0.45	-0.17
@ MCH (Mean Corpuscular Hb)	pg	30.20	28.35	-0.37	0.06
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.70	29.20	0.29	0.24
! Platelet Count	1000/ μ L	198.00	238.00	-2.27	-1.04
@ Hematocrit	%	32.80	26.60	-0.75	-0.69
@ RDW CV%	%	16.50	16.30	-1.13	-1.20

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam

PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A,

Kohinoor Mall, Kiroi Road, Kurla (W),

Mumbai - 400070

End of Report