

Analyte	Instrument	Result Value	Standard Unit
X pH	DIRUI H100	7.5	-
X Specific Gravity	DIRUI H100	1.005	-
X Protein (Urine)	DIRUI H100	2+	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**RIGHT CARE PATHOLOGY LAB (603)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 220207 and Sample No 01**

**Report Date : 31/08/2022**

**Instrument : DIRUI H100**

Analyte	Standard Unit	Result Value	Accepted Value
X pH	-	7.5	6,6.5,7
X Specific Gravity	-	1.005	1.015,1.020,1.025
@ Glucose (urine)	-	3+	+2,+3,+4
X Protein (Urine)	-	2+	+1,Neg,Tr
@ Ketone (urine)	-	Negative	+1,Neg,Tr
@ Urobilinogen (urine)	-	Normal	+1,Nor
@ Nitrite (urine)	-	Absent	Abs

**Instrument : Manual**

Analyte	Standard Unit	Result Value	Accepted Value
@ Blood (urine)	-	Negative	+1,Neg,Tr
@ Bilirubin (urine)	-	Negative	+1,Neg

Legend @ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	3
EQAS Score Clinical Path - Urine Routine Chemistry	66.67 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070

End of Report