

**RIGHT CARE PATHALOGY LAB**

RCP - 07

Proficiency Testing outliers and corrective action records

PAGE 1 OF 1

Sr. No	Date	Test Name	Lab Reported Result	Peer / Expected Value	Details of Root Cause analysis and Corrective action taken	Done by	Sign
1.	05-09-2022	Urine: PH	7.5	6,6.5,7	RCA: Checked for sample processing process, Reagent expiry, stability all factors were within acceptable limit. No specific deviation observed. Checked for Clinical correlation of the patient reported reports, it was also satisfactory. Sample reprocessing was not possible due to sample stability. Corrective Action: No specific corrective action taken, observe all abnormal reports for clinical correlation and observe EQAS reports in upcoming cycle.	Ms. Pranali	
2.		Urine: Specific Gravity	1.005	1.015,1.020,1.025			
3.		Urine: Protein	2+	+1,Neg,Tr			
4.							
5.							
6.							
7.							

Issue No. 1

Issue Date : 10-11-2021

Amend No. 00

Amend Date. 00

Page 1 of 1

Prepared & Approved By :
Ms. Jyoti Phadare