



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Jerath Test Lab & Allergy Testing Centre. 14 Link road, adjacent to SBI, Near, Amardas Chawk, Jalandhar, Punjab-144001	Certificate No.	CC223921000015235F
		Service request No & date	A/02/19/11/21
		Date of calibration	20/11/2021
		Valid up to	19/11/2022
		Date of issue	20/11/2021

Instrument Details			
Instrument name	Pipette	Model No.	----
Make	-----	Location	Lab
Range / Size	100-1000 μ l	Accuracy	-----
Least Count	5 μ l	Visual Inspection	OK
I.D. No.	PP-04		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weights	A Precision Balance	
Make	-----	AND	
Model / SR No.	-----	GH-252/----	
Certificate No.	21.05.001	21.02.09.001	
Calibration Validity	10/01/2023	15/02/2022	
Calibration by	Reliable Technical Services	Reliable Technical Services	

Environmental Condition	Temperature	25 \pm 4 $^{\circ}$ C	Calibration Reference
	Relative Humidity	50 \pm 20 %	Calibration Procedure

Calibration Results			
Serial No.	Std. Reading in (μ l)	Measured Reading in (μ l)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	100	99.98	±0.21 μ l
02.	200	199.56	
03.	400	399.53	
04.	600	599.47	
05.	800	799.29	
06	1000	999.19	

Remarks:

- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above result are valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(D. R. SAXENA)
Form No.-5.10 F-01



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

--- End of Report ---



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)

Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer name And Address	M/S: Jerath Test Lab & Allergy Testing Centre. 14 Link road, adjacent to SBI, Near, Amardas Chawk, Jalandhar, Punjab-144001	Certificate No.	CC223921000015234F
		Service request No & date	A/02/19/11/21
		Date of calibration	20/11/2021
		Valid up to	19/11/2022
		Date of issue	20/11/2021

Instrument Details			
Instrument name	Pipette	S. No.	-----
Make / Model No.	-----	Location	Lab
Range / Size	10-100 µl	Accuracy	-----
Least Count	1 µl	Visual Inspection	OK
I.D. No.	PP-05		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weights	A Precision Balance	
Make	-----	AND	
Model / SR No.	-----	GH-252/----	
Certificate No.	21.05.001	21.02.09.001	
Calibration Validity	10/01/2023	15/02/2022	
Calibration by	Reliable Technical Services	Reliable Technical Services	

Environmental Condition	Temperature	25±4°C	Calibration Reference
	Relative Humidity	50±20 %	Calibration Procedure

Calibration Results			
Serial No.	Std. Reading in (µl)	Measured Reading in (µl)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	10	10.02	±0.1 µl
02.	30	29.97	
03.	50	49.79	
04.	70	69.65	
05.	100	99.46	

Remarks:

❖ This result of calibration refers to only to the particular item submitted for calibration.
❖ The above result are valid at the time of and under the stated conditions measurement.
❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By
(Calibration Engg./TM)
(DIPSA XENA)
Form No.-5.10 F-01



Approved By
(QM/TM)
(NEERA TYAGI)
Page No. 1 of 1

--- End of Report---