



arkray healthcare Pvt.Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,
Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15
Facsimile: 0261-239 7719, URL <http://www.arkray.co.in>

Ref- AHPL/PT/2022-23
Date -10th Nov 2022



Certificate Of Calibration

- INSTRUMENT NAME : SPOTCHEM EL
- INSTRUMENT SERIAL NUMBER :12002035
- CALIBRATION DATE :10/11/2022
- DUE DATE OF CALIBRATION :09/11/2023

Self Check – Passed

- Check Plate results : Good
- Probe test :Good
- E plate movement :Good
- E plate results :Good

On the basis of above mentioned checks, this is to certify that SPOTCHEM EL 1520 installed at **MALAR DIAGNOSTICS , KELAMBAKKAM** has been calibrated on 10/11/2022 and performance of the instrument is satisfactory . This is valid upto 09/11/2023. If any major breakdown occurs, instrument must be re-evaluated for performance and calibration

<p>For, Arkray Healthcare Pvt Ltd.</p> <p><i>P. Sulthan Ahmed</i></p> <p>Sulthan Ahmed P_L Application specialist, Chennai</p>	<p></p> <p></p> <p>Customer's Signature and Stamp</p>
--	--