



CALIBRATION CERTIFICATE

Certificate No.	MES/21-22/THQHKT/REF/J01	Date of Issue : 15/01/2022
		Calibration Done On : 04/01/2022
		Calibration Due On : 03/01/2023

Customer Name	TALUK HEAD QUARTERS HOSPITAL, KODUNGALLUR, THRISSUR.
---------------	---

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	LG	512PRDE008202	-	LAB

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/SEP/21/I-45-01	19/09/2022
2	Digital Thermo Hygrometer	Equinox & EQ-308 CTH	20191602	ACS/JUN/21/I-32-01	17/06/2022

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
---------------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA

	Measured Reading	Range
Body	5.7	2 to 8 °C
Freezer	-13.5	-20 to -10 °C

Calibration Status	Pass	✓	Fail	
---------------------------	------	---	------	--

Calibrated By

(Signature)

Vivek Mahendran (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)