



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2022



PC-1024

Lab Name **JYOTI DENTAL & DIAGNOSTICS**

Lab No **16109**

Constituent Group **Chemistry I**

Date of Result Entered : **19/10/2022**

PT item **Lyophilized Serum**

Date of Report Published : **03/11/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Mindray	192	112.72	5.86	6.60	103.5 mg/dL	-1.40	0.95
2	UREA	Urease UV / GLDH	Mindray	235	129.02	7.49	9.66	133 mg/dL	0.41	1.26
3	CREATININE	Enzymatic Colorimetric	Mindray	71	8.60	6.54	0.56	8.2 mg/dL	-0.71	0.13
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Mindray	205	5.54	10.49	0.58	5.04 mg/dL	-0.86	0.08
5	T-PROTEIN	Biuret - Colorimetric	Mindray	244	5.48	7.03	0.38	5.5 g/dL	0.05	0.05
6	ALBUMIN	BCG - colorimetric	Mindray	228	3.36	7.21	0.24	3.2 g/dL	-0.66	0.03
7	CALCIUM	Arsenazo III	Mindray	192	10.84	7.15	0.78	10.5 mg/dL	-0.44	0.11
8	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Mindray	147	6.68	9.97	0.67	6.06 mg/dL	-0.93	0.11
9	URIC ACID	Enzymatic / Uricase Colorimetric	Mindray	241	6.06	8.60	0.52	6.1 mg/dL	0.08	0.07
10	CHOLESTEROL	CHOD-PAP	Mindray	249	110.37	6.54	7.22	111.6 mg/dL	0.17	0.91
11	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Mindray	238	136.98	7.93	10.87	135.3 mg/dL	-0.15	1.41
12	HDL	Direct method / Enzymatic colorimetric	Mindray	183	25.97	10.48	2.72	27.2 mg/dL	0.45	0.40
13	AST	UV kinetic(with & without PLP (P-5-P))	Mindray	222	37.45	9.82	3.68	38.2 U/L	0.20	0.49
14	ALT	UV kinetic(with & without PLP (P-5-P))	Mindray	211	309.33	16.40	50.72	339 U/L	0.58	6.98
15	ALP	PNP AMP kinetic	Mindray	168	230.13	14.46	33.27	195.9 U/L	-1.03	5.13

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

HEALTHHEART DIAGNOSTICS
HOUSE NO-7,HARDEV PURI,NATHU CHOWK
SHAHDARA DELHI-110093

Coordinator Contact Details:

Dr. Pamela Christudoss

Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******