



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2022



PC-1024

Lab Name **KUNNAMANGALAM NEETHI MEDICAL LABORATORY AND X-RAY CENTRE**

Lab No **16325**

Constituent Group **Chemistry I**

Date of Result Entered : **10/09/2022**

PT item **Lyophilized Serum**

Date of Report Published : **05/10/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Transasia / Erba	957	185.74	6.16	11.43	169 mg/dL	-1.46	0.74
2	UREA	Urease UV / GLDH	Transasia / Erba	894	100.59	7.45	7.49	86 mg/dL	-1.95	0.50
3	CREATININE	Enzymatic Colorimetric	Transasia / Erba	462	3.21	6.46	0.21	2.98 mg/dL	-1.11	0.02
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Transasia / Erba	925	3.03	9.61	0.29	2.25 mg/dL	-2.68	0.02
5	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	953	5.40	8.08	0.44	4.3 g/dL	-2.52	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	978	3.11	5.75	0.18	2.6 g/dL	-2.85	0.01
7	CALCIUM	Arsenazo III	Transasia / Erba	678	10.43	8.73	0.91	9.3 mg/dL	-1.24	0.07
8	URIC ACID	Enzymatic / Uricase Colorimetric	Transasia / Erba	974	7.17	8.97	0.64	6.4 mg/dL	-1.20	0.04
9	CHOLESTEROL	CHOD-PAP	Transasia / Erba	950	99.42	9.19	9.13	82 mg/dL	-1.91	0.59
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	923	156.54	8.66	13.56	144 mg/dL	-0.92	0.89
11	HDL	Direct method / Enzymatic colorimetric	Transasia / Erba	706	25.89	11.94	3.09	22 mg/dL	-1.26	0.23
12	SODIUM	ISE - Direct	Transasia / Erba	218	135.34	3.30	4.47	120.9 mmol/L	-3.23	0.60
13	POTASSIUM	ISE - Direct	Transasia / Erba	211	3.13	5.25	0.16	2.79 mmol/L	-2.07	0.02
14	CHLORIDE	ISE - Direct	Transasia / Erba	167	96.43	4.29	4.14	88.6 mmol/L	-1.89	0.64
15	AST	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	843	63.00	10.06	6.34	57 U/L	-0.95	0.44
16	ALT	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	831	86.68	14.46	12.54	76 U/L	-0.85	0.87
17	ALP	PNP AMP kinetic	Transasia / Erba	867	66.65	15.42	10.28	54 U/L	-1.23	0.70

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

**KUNNAMANGALAM NEETHI MEDICAL LABORATORY AND X-RAY CENTRE
KUNNAMANGALAM
KOZHIKODE
KERALA673571**

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******