



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2022



PC-1024

Lab Name **KUNNAMANGALAM NEETHI MEDICAL LABORATORY AND X-RAY CENTRE**

Lab No **16325**

Constituent Group **Chemistry I**

Date of Result Entered : **10/08/2022**

PT item **Lyophilized Serum**

Date of Report Published : **05/09/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Transasia / Erba	949	193.37	6.69	12.93	206 mg/dL	0.98	0.84
2	UREA	Urease UV / GLDH	Transasia / Erba	888	62.12	7.66	4.76	61 mg/dL	-0.24	0.32
3	CREATININE	Enzymatic Colorimetric	Transasia / Erba	465	2.54	7.18	0.18	2.64 mg/dL	0.55	0.02
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Transasia / Erba	942	3.62	10.23	0.37	3.21 mg/dL	-1.11	0.02
5	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	961	6.28	7.74	0.49	6.3 g/dL	0.04	0.03
6	ALBUMIN	BCG - colorimetric	Transasia / Erba	967	3.50	5.54	0.19	3.4 g/dL	-0.52	0.01
7	CALCIUM	Arsenazo III	Transasia / Erba	666	11.46	9.46	1.08	11.6 mg/dL	0.13	0.08
8	URIC ACID	Enzymatic / Uricase Colorimetric	Transasia / Erba	923	5.29	9.19	0.49	5.2 mg/dL	-0.19	0.03
9	CHOLESTEROL	CHOD-PAP	Transasia / Erba	929	114.16	8.16	9.32	107 mg/dL	-0.77	0.61
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	886	82.80	13.18	10.92	75 mg/dL	-0.71	0.73
11	HDL	Direct method / Enzymatic colorimetric	Transasia / Erba	713	28.02	10.91	3.06	27 mg/dL	-0.33	0.23
12	SODIUM	ISE - Direct	Transasia / Erba	208	139.95	2.87	4.01	140 mmol/L	0.01	0.56
13	POTASSIUM	ISE - Direct	Transasia / Erba	215	4.58	4.49	0.21	4.6 mmol/L	0.10	0.03
14	CHLORIDE	ISE - Direct	Transasia / Erba	170	102.85	4.27	4.39	109 mmol/L	1.40	0.67
15	AST	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	843	64.99	11.93	7.75	74 U/L	1.16	0.53
16	ALT	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	817	47.89	15.74	7.54	55 U/L	0.94	0.53
17	ALP	PNP AMP kinetic	Transasia / Erba	861	100.55	15.05	15.13	105 U/L	0.29	1.03

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

**KUNNAMANGALAM NEETHI MEDICAL LABORATORY AND X-RAY CENTRE
KUNNAMANGALAM
KOZHIKODE
KERALA673571**

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******