

VIEW LAB MONTHLY SUMMARY

Lab Name	RIA DIAGNOSTICS		Details About Robust Analysis							
Lab No	1863		Detail About Monthly Summary							
Month	October		Detail about SDI							
Year	2022									
Constituent Group	Chemistry I									
Click on the analyte to view Graphical Data			All Analyser Result		Print					
			Date of Result Entered : 14/10/2022							
			Date of Report Published : 03/11/2022							
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE	GOD-POD	Transasia / Erba	916	114.55	6.42	7.35	116 mg/dL	0.20	0.49
2	UREA	Urease UV / GLDH	Transasia / Erba	895	130.44	7.26	9.47	132 mg/dL	0.16	0.63
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Transasia / Erba	442	7.70	9.52	0.73	7.8 mg/dL	0.14	0.07
4	T-PROTEIN	Biuret - Colorimetric	Transasia / Erba	946	5.83	7.82	0.46	5.5 g/dL	-0.72	0.03
5	ALBUMIN	BCG - colorimetric	Transasia / Erba	951	3.31	5.80	0.19	3.3 g/dL	-0.05	0.01
6	CHOLESTEROL	CHOD-PAP	Transasia / Erba	918	105.66	8.20	8.66	103 mg/dL	-0.31	0.57
7	AST	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	844	38.01	14.25	5.42	42 U/L	0.74	0.37
8	ALT	UV kinetic(with & without PLP (P-5-P))	Transasia / Erba	819	276.92	16.40	45.41	261 U/L	-0.35	3.17
9	ALP	PNP AMP kinetic	Transasia / Erba	816	209.35	15.19	31.81	230 U/L	0.65	2.23
SDI Range		Interpretation								
Within -1.00 to +1.00		Excellent.								
Within ±1.01 to ±2.00		Good.								
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out