



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 806

Distribution No.: 157-B

Month/Year: July/2022

Instrument ID: TB-84002074

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 16-10-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	6	5.9	11.9	10.47	0.0280	2.00	0.1	0.1	0.0070	0.00
RBC x10 ⁶ /µl	1	4.71	4.68	9.39	8.75	0.0070	3.20	0.03	0.04	0.0020	-0.27
Hb g/dl	1	12.2	12.1	24.3	25.4	0.0200	-1.65	0.1	0.1	0.0070	0.00
HCT%	1	42.8	42.4	85.2	78.2	0.1260	1.89	0.4	0.4	0.0230	0.00
MCV-fl	1	90.9	90.5	181.4	178.7	0.2390	0.37	0.4	0.3	0.0200	0.34
MCH-Pg	1	25.9	25.9	51.8	58	0.0550	-4.18	0	0.2	0.0110	-0.90
MCHC-g/dl	1	28.6	28.5	57.1	64.8	0.1120	-2.31	0.1	0.3	0.0200	-0.67
Plt. x10 ³ /µl	1	100	97	197	214	1.05	-0.55	3	4	0.26	-0.19
Retic %	2	3	2.8	5.8	3.1	0.08	0.93	0.2	0.2	0.01	0.00

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=00 , Poly=70 L=16, E=02, Mono/Promono=02 , B1=10 P.M.=00, Mye=00, Meta=00, Other=	Poly: 42 - 56, Myelo: 13 - 23, Meta: 8 - 17, Lympho: 4 - 8, Eos: 2 - 5, Promyelo: 1-5, nRBC/Blast/Baso/Mono: 0 - 5		
RBC Morphology	3		Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis		
Diagnosis	3		Chronic Myeloid Leukemia (Chronic Phase)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 157--B	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 ³ /µl	1	373	366	83.61	91.53	7.92	3.28	8.47	5.19
RBC x10 ⁶ /µl	1	373	373	85.25	87.4	5.9	4.56	8.85	8.04
Hb g/dl	1	373	373	87.67	85.52	4.83	5.63	7.5	8.85
HCT%	1	373	365	89.59	89.59	7.12	5.21	3.29	5.2
MCV-fl	1	373	364	94.51	84.89	3.57	10.44	1.92	4.67
MCH-Pg	1	373	364	87.09	73.35	6.87	18.41	6.04	8.24
MCHC-g/dl	1	373	364	92.31	87.91	4.4	6.87	3.29	5.22
Plt. x10 ³ /µl	1	373	364	91.21	93.13	5.22	3.3	3.57	3.57
ReticCount%	2	373	273	91.21	95.6	5.49	1.47	3.3	2.93
PS Assessment	3	373	349	Satisfactory :90.89%, Borderline Sat. :6.97%, Unsatisfactory :2.14%					

Comments:

1). Among Lab (EQA) : CBC result for RBC & MCH unacceptable, please check calibration/human error.PS Diagnosis not reported, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----