



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments

304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechs.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL THERMO-HYGROMETER

Calibration Certificate No.	22100014135	Issued On	06/09/2022
ULR No.	CC214622100014135F	Calibration Date	03/09/2022
Job Identification No.	ICS/C/T/09/14135	Next Due Date	02/09/2023
Ref. No.	SRF- Dated: 30/08/2022		

CALIBRATED FOR: Dr. Khanna's Pathcare Pvt. Ltd.
C/o Sanjeevan Medical Research Centre Pvt. Ltd.
24, Ansari Road,
Daryaganj, New Delhi-110002.

EQUIPMENT DESCRIPTION

Name	Digital Thermo-Hygrometer		
Make/Model	HTC-1	Visual Inspection	Ok
Serial No.	---	I.D. No.	SMRC/GEN/LAB/INSA/LAB-01
Range	Up to 60°C / Up to 99 % (RH)	Least Count	0.1°C / 1 % RH
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Yudian	ETL/2113/22/C/F/061	FCRI, Palakkad	02-03-2022 to 04-03-2022	04-03-2023
Temp. & Hum. Ind. with Sensor	Rotronic	ETL/1703/21/C/F/506	FCRI, Kerala	30/12/2021 - 04-01-2022	04/01/2023

USED EQUIPMENT DETAILS: Environmental Chamber

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-TH06 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration
(On Working Range)

U.U.C. Reading		Corrected Value of Standard (Average)		Uncertainty at approx 95% C.L and coverage factor k=2	
Temperature	Humidity	Temperature	Humidity	Temperature	Humidity
15.0 °C	20 % RH	15.062 °C	20.51 % RH	± 0.33°C	± 1.6 % RH
20.0 °C	40 % RH	20.099 °C	40.92 % RH	± 0.33°C	± 1.6 % RH
30.0 °C	75 % RH	30.131 °C	76.25 % RH	± 0.33°C	± 1.6 % RH
40.0 °C	90 % RH	40.172 °C	91.69 % RH	± 0.33°C	± 1.6 % RH

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by  RAJESH KAPOOR
Calibration Head

Calibration Head