



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIAL PRESSURE GAUGE

Certificate No:	22300007715	Issued On:	06/09/2022
ULR No:	CC214622300007715F	Calibration Date:	03/09/2022
Job Identification No:	ICS/C/VP/08/7715	Next Calibration Date:	02/09/2023
Ref. No:	SRF- Dated: 30/08/2022		

CALIBRATED FOR: Dr. Khanna's Pathcare Pvt. Ltd.
C/o Sanjeevan Medical Research Centre Pvt. Ltd.
24, Ansari Road,
Daryaganj, New Delhi-110002.

EQUIPMENT DESCRIPTION

Name	Dial Pressure Gauge,		
Make/Model	Shiva	Visual Inspection	Ok
Serial No.	---	I.D. No.	---
Range	0 to 2 Kg/cm ²	Least Count	0.05 Kg/cm ²
Location	---	Calibration Site	In Lab.

ENVIRONMENTAL CONDITIONS

Temperature	20.0 °C ± 2.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Pressure Gauge	Vijay Enterprise	PSL/0496/0622/ P/0904	FCRI,PALAKKAD	24/06/2022	24/06/2023

USED EQUIPMENT DETAILS Pressure Comparator
PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-P02, DKDR-6/1 (Comparison Method)

RESULTS: Mechanical Calibration

Parameter	UUC Pressure in (kg/cm ²)	UUC Pressure in (bar)	Indicated Values on Standard in (bar) (Avg.)	Uncertainty At Approx 95% C.L And Coverage Factor K=2
Pressure	0.00	0.0000	0.0000	± 0.03 bar
	0.50	0.4903	0.4908	
	1.00	0.9807	0.9819	
	1.50	1.4710	1.4727	
	2.00	1.9613	1.9639	


1kg/cm² = 0.980665bar

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.
- *** END OF REPORT ***

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by 
Authorised Signatory
RAJESH KAPOOR
ration-Head