

H 560 – 3 Part Haematology Analyzer

**INSTALLATION QUALIFICATION
OPERATIONAL QUALIFICATION
PERFORMANCE QUALIFICATION
(IQ/OQ/PQ)**

**HEALTHHEART DIAGNOSTICS
BUILDING-07,HARDEV PURI,
NATHU CHOWK,SHAHDARA,
DELHI-110093**

By:- Avantor Performance Materials India Limited

CERTIFICATE OF INSTALLATION

INSTRUMENT NAME : H 560
SERIAL NUMBER : K1104B2146041
CUSTOMER NAME : HEALTHHEART DIAGNOSTICS
**Address : BUILDING-07, HARDEV PURI,
NATHU CHOWK, SHAHDARA,
DELHI-110093**

The undersigned performers certify that the Installation Qualification Protocol has been successfully completed for the instrument stated above.

Engineer

Signature :

Name : SANJEEV
Designation : Regional Service Manager

Company : Avantor Performance Materials India Limited

Quality Manager

Signature :

Name : DR.POOJA

Designation : DIECTOR

Organization : HEALTHHEART DIAGNOSTICS

INSTALLATION REPORT

Customer Name : HEALTHHEART DIAGNOSTICS

Place : DELHI

Department : Pathology / Laboratory

Contact Person : DR.POOJA

Instrument Name : H 560 (3 Part Hematology Analyzer)

Serial Number : K1104B2146041

Date of Installation : 23.06.2022



The Analyzer was installed along with the necessary standard accessories. All basic requirements and adjustments were checked and are found to be satisfactory. The preliminary standardization of the analyzer and the training were provided to the complete satisfaction of the customer. The analyzer is found to be working satisfactorily.

For Avantor Performance Materials India Limited

**Nasim Ahmad
Regional Manager
Lucknow**

INSTALLATION QUALIFICATION

IQ PROTOCOL

- 1. System Unpacking*
- 2. System Checking For any Damages*
- 3. All accessories as per check List verification*
- 4. Space Requirement*
- 5. Power Requirement*
- 6. Reagents, Control, Calibrators verification.*

1. SYSTEM UNPACKING & SYSTEM CHECKING

Avantor Performance Materials India Ltd

Registered Office: 17th Floor, Building No 5, DLF Cyber City, Phase III, Gurugram -122002, Haryana, India

H 560 is unpacked and checked for physical Damages and found to be in good condition.

2. ACCESSORIES CHECK-

Check all accessories as per check List

3. SPACE REQUIREMENT –

- **Checked site for proper space allocation. At least 100 CM on each side and enough room on below the countertop to accommodate the instrument unit and waste & Reagents.**

| DIMENTIONS | |
|------------|-------------|
| PARAMETER | Analyzer OV |
| Width mm | 320 |
| Depth mm | 400 |
| Height mm | 410 |
| Weight Kg | <=24 |

4. POWER REQUIREMENT –

| Requirement | Acceptable Range | Observed Ranged |
|---------------------|------------------|-----------------|
| Input Voltage | 100-240V~ | 225 V |
| Line Frequency | 50 HZ | 50 HZ |
| Ambient Temperature | 15 C – 30 C | 23 |
| Grounding Voltage | 0 to 5 V | 3 v |

5. Control, Calibrators/Standards Check-

| NAME | Status | Remark |
|----------------------|--------|--------|
| Controls | YES | OK |
| Calibrator/Standards | YES | OK |

OPERATIONAL QUALIFICATION

The Operational Qualification procedure specifies the methodology for installation of specified system after successful installation qualification. Successful completion of procedure identifies that system is ready for operation and subsequent performance analysis.

OO PROTOCOL

1. System Connections
2. Waste Connections
3. Reagent connections
4. Customizing the Analyzer Software
5. System Booting , Initialization & Check
6. Maintenance procedures.
7. Customer Training – Operation & Maintenance.

1. SYSTEM CONNECTION –

| SYSTEM CONNECTIONS | | | |
|--------------------|--------------------|--------------------------------|--------|
| S/N | NAME | Check | Remark |
| 1 | Input Power supply | Connect, Operator manual 4.2.2 | OK |
| 2 | System setting | operating manual 9.2 | OK |
| 3 | Reagent Setup | operating manual 9.2.6 | OK |
| 4 | Parameter Setting | operating manual 9.2.4 | OK |

2. REAGENT & WASTE CONNECTIONS -

| REAGENT AND WASTE CONNECTION CHECK | | |
|------------------------------------|---------------------------------|--------|
| NAME | CHECK | Remark |
| Reagent Connection | Operation Manual 4.3, and 9.2.6 | OK |
| Waste Connection | Operating Manual 4.3 | OK |

3. CUSTOMIZING ANALYZER SOFTWARE –

| CUSTOMIZING ANALYZER SOFTWARE | | | |
|-------------------------------|-------------------|------------------------|--------|
| S/N | Parameters | Check | Remark |
| 1 | Analyzing Unit | Operating Manual 2.1 | OK |
| 2 | Startup and Logon | Operating Manual 5.3 | OK |
| 3 | QC and Control | Operating Manual 7.1 | OK |
| 4 | Sample Analysis | Operating Manual 5.5.2 | OK |
| 5 | Shutdown | Operating Manual 5.7 | OK |

4. SYSTEM BOOTING, INITIALIZATION & Check -

**Power ON System and check instrument Initialization successfully.
Check following Parameters.**

| STARTUP | | |
|----------------|-------------------|---------------|
| S/N | Parameters | Remark |
| 1 | Reaction chamber | Ok |
| 2 | Ambience | Ok |
| 3 | Waste | Ok |
| 4 | Diluent | Ok |
| 5 | LH Lyse | Ok |
| 6 | Diff Lyse | Ok |
| 7 | Probe Cleaner | Ok |
| 8 | Control | Ok |

MAINTENANCE PROCEDURES:-

| MAINTENANCE | | | |
|--------------------|-------------------|-------------------------|---------------|
| S/N | Parameters | Check | Remark |
| 1 | Maintenance | Operation Manual 10.2.1 | OK |
| 2 | Troubleshooting | Operation Manual 11.1 | OK |

PERFORMANCE QUALIFICATION

The Performance of system is to be verified by below mentioned tests The results should be with in range as per the instruments and control sheet.

PQ Protocol:-

1. Background check
2. Carry over check
3. Reproducibility Check
4. Sample Precision, Accuracy and Performance test
5. Control Check (Three Levels)

1. Background check

| BACKGROUND CHECK TEST | | | | | | |
|-----------------------|------------|------|------|------|------|-------|
| | Date/Time | WBC | RBC | HGB | PLT | WBC-D |
| 10 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 9 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 8 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 7 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 6 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 5 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 4 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 3 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 2 | 28/07/2022 | 0 | 0 | 0 | 0 | 0 |
| 1 | 28/07/2022 | 0.04 | 0 | 0 | 0 | 0 |
| Limit | | 0.1 | 0.02 | 0.1 | 5 | 0.2 |
| Result | | Pass | Pass | Pass | Pass | Pass |

2. Carry Over Check.

| CARRYOVER CHECK TEST | | | | | | | |
|-----------------------------|------------|-------|-------|-------|-------|--------|-------|
| Substance | Control | | | | | | |
| | Date/Time | WBC | RBC | HGB | HCT | PLT | WBC-D |
| High-Level 1 | 28/07/2022 | 21.06 | 5.87 | 17.7 | 63.6 | 467 | 21.78 |
| High-Level 2 | 28/07/2022 | 20.99 | 5.8 | 17.4 | 63 | 463 | 21.28 |
| High-Level 3 | 28/07/2022 | 20.69 | 5.81 | 17.5 | 62.8 | 458 | 21.96 |
| Low-Level 1 | 28/07/2022 | 3.83 | 2.24 | 5.8 | 22.5 | 57 | 3.82 |
| Low-Level 2 | 28/07/2022 | 3.7 | 2.23 | 5.7 | 22.3 | 61 | 3.8 |
| Low-Level 3 | 28/07/2022 | 3.67 | 2.23 | 5.7 | 22.3 | 65 | 3.8 |
| Carryover% | | 0.90% | 0.30% | 0.80% | 0.50% | -2.00% | 0.10% |
| Limit% | | ≤1.0% | ≤1.0% | ≤1.0% | ≤1.0% | ≤1.0% | ≤1.0% |
| Result | | Pass | Pass | Pass | Pass | Pass | Pass |

3. Reproducibility Check

| Reproducibility Check | | | | | | |
|------------------------------|-------------|------|------|------|-------|-----|
| Date/Time | Run time | WBC | RBC | HGB | MCV | PLT |
| 28/07/2022 | 1 | 7.29 | 4.69 | 13.1 | 101 | 223 |
| 28/07/2022 | 2 | 7.16 | 4.7 | 13.1 | 100.9 | 227 |
| 28/07/2022 | 3 | 7.11 | 4.66 | 13.1 | 101.3 | 226 |
| 28/07/2022 | 4 | 7.24 | 4.6 | 12.9 | 101.2 | 226 |
| 28/07/2022 | 5 | 6.94 | 4.69 | 13.1 | 100.9 | 223 |
| 28/07/2022 | 6 | 7.36 | 4.7 | 13.1 | 101 | 219 |
| 28/07/2022 | 7 | 7.34 | 4.68 | 13 | 101.1 | 228 |
| 28/07/2022 | 8 | 7.4 | 4.74 | 13.2 | 100.9 | 231 |
| 28/07/2022 | 9 | 7.48 | 4.74 | 13.2 | 101.1 | 222 |
| 28/07/2022 | 10 | 7.16 | 4.77 | 13.2 | 101.3 | 220 |
| 28/07/2022 | 11 | 7.23 | 4.73 | 13.3 | 101.7 | 237 |
| 28/07/2022 | Mean | 7.25 | 4.71 | 13.1 | 101.1 | 226 |

| | | | | | | |
|------------|-----------------|------|-------|------|-------|------|
| 28/07/2022 | SD | 0.16 | 0.034 | 0.08 | 0.25 | 5.5 |
| 28/07/2022 | Min | 6.94 | 4.66 | 13.1 | 100.9 | 219 |
| 28/07/2022 | Max | 7.48 | 4.77 | 13.3 | 101.7 | 237 |
| 28/07/2022 | R | 0.54 | 0.11 | 0.3 | 0.8 | 18 |
| 28/07/2022 | d(Min) | 0.02 | 0.01 | 0 | 0 | 0 |
| 28/07/2022 | d(Max) | 0.31 | 0.06 | 0.2 | 0.6 | 11 |
| | CV% | 2.2 | 0.7 | 0.6 | 0.3 | 2.4 |
| | Limit(%) | 2.5 | 1.5 | 1 | 1 | 4 |
| | Result | Pass | Pass | Pass | Pass | Pass |

| Reproducibility Check | | | | | | | |
|------------------------------|-----------------|-------|-------|-------|-------|------|-------|
| Date/Time | Run time | Neu# | Lym# | Mon# | Eos# | Bas# | IMG# |
| 28/07/2022 | 1 | 4.14 | 1.48 | 0.24 | 1.42 | 0.01 | 0.02 |
| 28/07/2022 | 2 | 4.14 | 1.41 | 0.23 | 1.37 | 0.01 | 0.01 |
| 28/07/2022 | 3 | 4.1 | 1.4 | 0.24 | 1.36 | 0.01 | 0.02 |
| 28/07/2022 | 4 | 4.25 | 1.41 | 0.24 | 1.33 | 0.01 | 0.03 |
| 28/07/2022 | 5 | 3.99 | 1.38 | 0.23 | 1.33 | 0.01 | 0.02 |
| 28/07/2022 | 6 | 4.34 | 1.41 | 0.25 | 1.35 | 0.01 | 0.02 |
| 28/07/2022 | 7 | 4.28 | 1.48 | 0.23 | 1.34 | 0.01 | 0.03 |
| 28/07/2022 | 8 | 4.29 | 1.52 | 0.21 | 1.37 | 0.01 | 0.03 |
| 28/07/2022 | 9 | 4.18 | 1.58 | 0.26 | 1.45 | 0.01 | 0.02 |
| 28/07/2022 | 10 | 4.18 | 1.39 | 0.24 | 1.34 | 0.01 | 0.02 |
| 28/07/2022 | 11 | 4.08 | 1.47 | 0.29 | 1.38 | 0.01 | 0.03 |
| 28/07/2022 | Mean | 4.17 | 1.45 | 0.24 | 1.37 | 0.01 | 0.02 |
| 28/07/2022 | SD | 0.107 | 0.065 | 0.021 | 0.038 | 0 | 0.006 |
| 28/07/2022 | Min | 3.99 | 1.38 | 0.21 | 1.33 | 0.01 | 0.01 |
| 28/07/2022 | Max | 4.34 | 1.58 | 0.29 | 1.45 | 0.01 | 0.03 |
| 28/07/2022 | R | 0.35 | 0.2 | 0.08 | 0.12 | 0 | 0.02 |
| 28/07/2022 | d(Min) | 0.01 | 0.02 | 0 | 0 | 0 | 0 |
| 28/07/2022 | d(Max) | 0.18 | 0.13 | 0.05 | 0.08 | 0 | 0.01 |
| | CV% | 2.6 | 4.5 | 8.9 | 2.8 | 0 | 28.7 |
| | Limit(%) | 6 | 6 | 16 | 20 | 30 | 25 |

| | | | | | | | |
|--|---------------|------|------|------|------|------|------|
| | Result | Pass | Pass | Pass | Pass | Pass | Pass |
|--|---------------|------|------|------|------|------|------|

| Reproducibility Check | | | | | | | |
|------------------------------|-------------|------|------|------|------|------|------|
| Date/Time | Run time | Neu% | Lym% | Mon% | Eos% | Bas% | IMG% |
| 28/07/2022 | 1 | 56.9 | 20.2 | 3.3 | 19.4 | 0.2 | 0.3 |
| 28/07/2022 | 2 | 57.7 | 19.7 | 3.3 | 19.1 | 0.2 | 0.2 |
| 28/07/2022 | 3 | 57.8 | 19.6 | 3.4 | 19.1 | 0.1 | 0.3 |
| 28/07/2022 | 4 | 58.7 | 19.5 | 3.4 | 18.3 | 0.1 | 0.4 |
| 28/07/2022 | 5 | 57.6 | 19.8 | 3.3 | 19.1 | 0.2 | 0.3 |
| 28/07/2022 | 6 | 59.1 | 19.1 | 3.4 | 18.3 | 0.1 | 0.3 |
| 28/07/2022 | 7 | 58.3 | 20.2 | 3.1 | 18.2 | 0.2 | 0.4 |
| 28/07/2022 | 8 | 58.1 | 20.5 | 2.8 | 18.5 | 0.1 | 0.4 |
| 28/07/2022 | 9 | 55.9 | 21.1 | 3.5 | 19.4 | 0.1 | 0.3 |
| 28/07/2022 | 10 | 58.3 | 19.4 | 3.4 | 18.8 | 0.1 | 0.2 |
| 28/07/2022 | 11 | 56.5 | 20.3 | 4 | 19.1 | 0.1 | 0.5 |
| 28/07/2022 | Mean | 57.6 | 20 | 3.3 | 18.9 | 0.1 | 0.3 |
| 28/07/2022 | SD | 0.95 | 0.59 | 0.3 | 0.43 | 0.05 | 0.09 |

| | | | | | | | |
|------------|-----------------|------|------|------|------|------|------|
| 28/07/2022 | Min | 55.9 | 19.1 | 2.8 | 18.2 | 0.1 | 0.2 |
| 28/07/2022 | Max | 59.1 | 21.1 | 4 | 19.4 | 0.2 | 0.5 |
| 28/07/2022 | R | 3.2 | 2 | 1.2 | 1.2 | 0.1 | 0.3 |
| 28/07/2022 | d(Min) | 0 | 0.2 | 0 | 0.1 | 0 | 0 |
| 28/07/2022 | d(Max) | 1.7 | 1.1 | 0.7 | 0.7 | 0.1 | 0.2 |
| | CV% | 1.6 | 2.9 | 9 | 2.3 | 36.9 | 28.7 |
| | Limit(%) | 6 | 6 | 16 | 20 | 30 | 25 |
| | Result | Pass | Pass | Pass | Pass | Pass | Pass |

| Reproducibility Check | | | | | | | |
|------------------------------|----------|------|------|------|--------|--------|------|
| Date/Time | Run time | HCT | MCH | MCHC | RDW-CV | RDW-SD | MPV |
| 28/07/2022 | 1 | 47.3 | 28 | 27.7 | 16.5 | 56.4 | 10.1 |
| 28/07/2022 | 2 | 47.4 | 27.9 | 27.7 | 16.3 | 55.7 | 9.8 |
| 28/07/2022 | 3 | 47.2 | 28 | 27.7 | 16.3 | 55.6 | 9.9 |
| 28/07/2022 | 4 | 46.6 | 27.9 | 27.6 | 16.3 | 55.9 | 10.1 |
| 28/07/2022 | 5 | 47.3 | 27.9 | 27.6 | 16.7 | 56.8 | 9.8 |
| 28/07/2022 | 6 | 47.5 | 27.9 | 27.6 | 16.3 | 55.5 | 9.9 |

| | | | | | | | |
|------------|-----------------|------|------|------|------|------|------|
| 28/07/2022 | 7 | 47.3 | 27.9 | 27.6 | 16.3 | 55.7 | 9.9 |
| 28/07/2022 | 8 | 47.8 | 27.8 | 27.6 | 16.2 | 55.3 | 10.1 |
| 28/07/2022 | 9 | 48 | 27.8 | 27.5 | 16 | 54.6 | 10.1 |
| 28/07/2022 | 10 | 48.4 | 27.7 | 27.3 | 16.1 | 55 | 9.8 |
| 28/07/2022 | 11 | 48.1 | 28.1 | 27.7 | 16.3 | 55.8 | 10 |
| 28/07/2022 | Mean | 47.6 | 27.9 | 27.6 | 16.3 | 55.6 | 9.9 |
| 28/07/2022 | SD | 0.42 | 0.12 | 0.12 | 0.19 | 0.63 | 0.13 |
| 28/07/2022 | Min | 47.2 | 27.7 | 27.3 | 16 | 54.6 | 9.8 |
| 28/07/2022 | Max | 48.4 | 28.1 | 27.7 | 16.7 | 56.8 | 10.1 |
| 28/07/2022 | R | 1.2 | 0.4 | 0.4 | 0.7 | 2.2 | 0.3 |
| 28/07/2022 | d(Min) | 0.1 | 0 | 0 | 0 | 0 | 0 |
| 28/07/2022 | d(Max) | 0.8 | 0.2 | 0.3 | 0.4 | 1.2 | 0.2 |
| | CV% | 0.9 | 0.4 | 0.4 | 1.2 | 1.1 | 1.3 |
| | Limit(%) | 1.5 | 1.5 | 1.5 | 2 | 2 | 3 |
| | Result | Pass | Pass | Pass | Pass | Pass | Pass |

Reproducibility Check

Avantor Performance Materials India Ltd

Registered Office: 17th Floor, Building No 5, DLF Cyber City, Phase III, Gurugram -122002, Haryana, India

| Date/Time | Run time | PDW | PCT | P-LCC | P-LCR | NRBC# | NRBC% |
|------------|-----------------|------|--------|-------|-------|--------|-------|
| 28/07/2022 | 1 | 16.4 | 0.225 | 60 | 26.6 | 0 | 0 |
| 28/07/2022 | 2 | 16.4 | 0.222 | 60 | 26.3 | 0 | 0 |
| 28/07/2022 | 3 | 16.4 | 0.224 | 59 | 25.9 | 0.002 | 0.02 |
| 28/07/2022 | 4 | 16.7 | 0.228 | 61 | 27.3 | 0.005 | 0.06 |
| 28/07/2022 | 5 | 16.4 | 0.219 | 59 | 26.4 | 0.005 | 0.06 |
| 28/07/2022 | 6 | 16.6 | 0.217 | 57 | 26.1 | 0.002 | 0.02 |
| 28/07/2022 | 7 | 16.5 | 0.226 | 60 | 26.1 | 0.002 | 0.02 |
| 28/07/2022 | 8 | 16.9 | 0.234 | 66 | 28.3 | 0.003 | 0.04 |
| 28/07/2022 | 9 | 16.6 | 0.223 | 62 | 27.8 | 0.003 | 0.04 |
| 28/07/2022 | 10 | 16.4 | 0.217 | 57 | 25.9 | 0.002 | 0.02 |
| 28/07/2022 | 11 | 16.5 | 0.238 | 65 | 27.5 | 0.002 | 0.02 |
| 28/07/2022 | Mean | 16.5 | 0.224 | 61 | 26.7 | 0.002 | 0.02 |
| 28/07/2022 | SD | 0.16 | 0.0069 | 3 | 0.86 | 0.0014 | 0.018 |
| 28/07/2022 | Min | 16.4 | 0.217 | 57 | 25.9 | 0 | 0 |
| 28/07/2022 | Max | 16.9 | 0.238 | 66 | 28.3 | 0.005 | 0.06 |
| 28/07/2022 | R | 0.5 | 0.021 | 9 | 2.4 | 0.005 | 0.06 |
| 28/07/2022 | d(Min) | 0 | 0 | 1 | 0.1 | 0 | 0 |
| 28/07/2022 | d(Max) | 0.4 | 0.014 | 6 | 1.6 | 0.003 | 0.04 |
| | CV% | 1 | 3.1 | 5 | 3.2 | 69 | 76.6 |
| | Limit(%) | 10 | 5 | 15 | 15 | 20 | 20 |
| | Result | Pass | Pass | Pass | Pass | Pass | Pass |

4. Sample Precision, Accuracy and Performance test

Nine tests of a sample in different time interval in two days have been run for the parameters and a statistical analysis of the data. The results are within defined range & with following precision & accuracy limits based on the reference range given by the manufacturer.

| | | PERFORMANCE QUALIFICATION PERFORMED FOR MINDRAY H 560HEMOTOLOGY WITH PATIENT SAMPLE | | | | | | | |
|---------|-------------------|---|---------|-------|------|------|-------|--------|-------|
| | | RUN | TIME | HB | WBC | RBC | HCT | PLT | MCV |
| DAY - 1 | DATE : 21/07/2022 | RUN 1 | 9:44 AM | 12.4 | 7.68 | 4.23 | 34.4 | 277 | 81.4 |
| | | RUN 2 | 9:44 AM | 12.4 | 7.9 | 4.21 | 34.3 | 273 | 81.6 |
| | | RUN 3 | 09.51am | 12.4 | 7.67 | 4.21 | 34.4 | 276 | 81.8 |
| | | RUN 4 | 09.51am | 12.4 | 7.76 | 4.24 | 34.9 | 268 | 82.4 |
| | | RUN 5 | 12.47pm | 12.4 | 7.76 | 4.26 | 35.5 | 272 | 82.5 |
| | | RUN 6 | 12.51pm | 12.4 | 7.74 | 4.21 | 34.7 | 268 | 82.4 |
| | | RUN 7 | 03.40pm | 12.4 | 7.77 | 4.15 | 33.6 | 271 | 81.1 |
| | | RUN 8 | 03.43pm | 12.5 | 7.79 | 4.21 | 34.5 | 274 | 81.9 |
| | | RUN 9 | 3.46pm | 12.5 | 7.86 | 4.27 | 35 | 286 | 81.1 |
| | | | | | | | | | |
| | | MEAN | | 12.42 | 7.77 | 4.22 | 34.59 | 273.89 | 81.80 |
| | | SD | | 0.04 | 0.07 | 0.04 | 0.53 | 5.51 | 0.55 |
| | | CV% | | 0.35 | 0.96 | 0.83 | 1.53 | 2.01 | 0.67 |
| Result | | | Pass | Pass | Pass | Pass | Pass | Pass | |

| | | PERFORMANCE QUALIFICATION PERFORMED FOR MINDRAY H 560HEMOTOLOGY WITH PATIENT SAMPLE | | | | | | | |
|----------------|--------------------------|--|---------|--------------|-------------|-------------|--------------|---------------|--------------|
| | | RUN | TIME | HB | WBC | RBC | HCT | PLT | MCV |
| DAY - 2 | DATE : 22/07/2022 | RUN 1 | 9:50 AM | 12.4 | 7.75 | 4.22 | 34.7 | 278 | 82.2 |
| | | RUN 2 | 09.53am | 12.4 | 7.64 | 4.18 | 34.6 | 261 | 82.7 |
| | | RUN 3 | 09.56am | 12.3 | 8.05 | 4.09 | 33.7 | 271 | 82.5 |
| | | RUN 4 | 12.56pm | 12.4 | 7.67 | 4.2 | 34.5 | 272 | 82 |
| | | RUN 5 | 12.59pm | 12.4 | 7.74 | 4.19 | 34.4 | 277 | 82.1 |
| | | RUN 6 | 13.02pm | 12.4 | 7.79 | 4.19 | 34.4 | 267 | 82.2 |
| | | RUN 7 | 03.20pm | 12.4 | 7.79 | 4.24 | 34.8 | 268 | 82 |
| | | RUN 8 | 03.23pm | 12.4 | 7.62 | 4.21 | 34.7 | 264 | 82.4 |
| | | RUN 9 | 3.26pm | 12.3 | 7.68 | 4.14 | 34.1 | 267 | 82.4 |
| | | | | | | | | | |
| | | MEAN | | 12.38 | 7.75 | 4.18 | 34.43 | 269.44 | 82.28 |
| | | SD | | 0.04 | 0.13 | 0.05 | 0.35 | 5.64 | 0.24 |
| | | CV% | | 0.36 | 1.67 | 1.08 | 1.01 | 2.09 | 0.29 |
| | | Result | | | Pass | Pass | Pass | Pass | Pass |

5. Control Check

| Name | Lot no | Expiry | | Result |
|---------------|---------------|---------------|--|---------------|
| BC 5D Level-L | BC1901BL | 10.09.2022 | | Pass |
| BC 5D level-N | BC1901BN | 10.09.2022 | | Pass |
| BC 5D level-H | BC1901BH | 10.09.2022 | | Pass |

CALIBRATION CERTIFICATE

NAME OF THE INSTRUMENT : H560

INSTRUMENT SL. NO. : K1104B2146041

USER NAME : HEALTHHEART DIAGNOSTICS

ADDRESS : BUILDING NO-07, HARDEV PURI, NATHU CHOWK,
SHAHDARA DELHI-110093

DATE OF CALIBRATION : 01st August 2022

CALIBRATOR LOT NO. : SC350

CALIBRATION:

- Calibration must be performed initially at installation and then every 6 months thereafter.
- Interim recalibration must be performed when
 - The instrument is moved to a different location
 - After replacing any component related to the process of dilution or measurement
 - When QC results indicate recalibration may be needed
 - When instructed by technical support to recalibrate as part of troubleshooting

The package insert which comes with the calibrator contains the target values to be used in calibrating. This insert must be saved per local regulations.

REPORTABLE RANGES (ANALYTICAL MEASUREMENT RANGES)

The manufacturer's claims for Reportable Range are as follows:

| Parameter | Reportable Range (AMR) Low | Reportable Range (AMR) High |
|----------------------------|-------------------------------|--------------------------------|
| WBC ($10^3/\mu\text{l}$) | 0.95 | 83.45 |
| HGB (g/dl) | 1.4 | 23.7 |
| RBC ($10^6/\mu\text{l}$) | 0.44 | 7.74 |
| PLT ($10^3/\mu\text{l}$) | 11 | 975 |

CALIBRATION DATA

| | Select | WBC | RBC | HGB | MCV | PLT |
|--------------|--------|-------|--------|-------|-------|--------|
| Target | | 9.60 | 4.50 | 13.7 | 88.9 | 259 |
| 1 | Yes | 9.9 | 4.49 | 13.9 | 88.9 | 255 |
| 2 | Yes | 9.8 | 4.47 | 13.6 | 88.9 | 251 |
| 3 | Yes | 9.7 | 4.46 | 13.8 | 89.6 | 241 |
| 4 | Yes | 9.8 | 4.47 | 13.8 | 89.7 | 247 |
| 5 | Yes | 9.9 | 4.42 | 13.6 | 89.6 | 236 |
| 6 | Yes | 9.6 | 4.48 | 13.7 | 89.5 | 245 |
| 7 | Yes | 9.7 | 4.47 | 13.5 | 89.7 | 246 |
| 8 | Yes | 9.9 | 4.48 | 13.8 | 89.7 | 242 |
| 9 | Yes | 9.8 | 4.45 | 13.8 | 89.6 | 243 |
| 10 | Yes | 9.7 | 4.46 | 13.9 | 89.6 | 243 |
| Mean | | 9.8 | 4.47 | 13.7 | 89.7 | 245 |
| CV% | | 1.1 | 0.4 | 0.9 | 0.2 | 2.2 |
| New Factor % | | 97.98 | 100.99 | 99.77 | 99.13 | 105.77 |
| Old Factor % | | 97 | 100 | 94 | 100 | 100 |

CALIBRATION FACTOR

| Parameter | Calibration Factor (%) | Date |
|-----------|--------------------------|-----------|
| WBC | 97.98 | 1/08/2022 |
| RBC | 100.99 | 1/08/2022 |
| HGB | 99.77 | 1/08/2022 |
| MCV | 99.13 | 1/08/2022 |
| PLT | 105.77 | 1/08/2022 |

This Certificate is valid up to July 2023



(Sanjeev Gupta)

Manager – Technical Services