

VIEW LAB MONTHLY SUMMARY

Lab Name	Bethel X - Ray , ECG & Clinical Laboratory	Details About Robust Analysis
Lab No	11638	Detail About Monthly Summary
Month	September	Detail about SDI
Year	2022	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/09/2022

Date of Report Published : 05/10/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	886	188.05	9.72	18.27	147.8 mg/dL	-2.20	1.23
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	856	51.66	12.91	6.67	62.5 mg/dL	1.63	0.46
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	562	2.62	10.95	0.29	2.73 mg/dL	0.38	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	827	2.94	13.86	0.41	3.45 mg/dL	1.25	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	690	5.72	8.35	0.48	5.75 g/dL	0.06	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	714	3.33	9.02	0.30	3.67 g/dL	1.13	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	782	5.67	13.33	0.76	6.83 mg/dL	1.53	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	789	109.84	10.64	11.68	117 mg/dL	0.61	0.83
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	764	161.37	11.70	18.88	181.2 mg/dL	1.05	1.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out