



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - OCTOBER 2022



PC 1024

Lab Name **SRI AMMAN MICRO LAB**
Constituent Group **Chemistry II**
PT Item **Lyophilized Serum**

Lab No **16124**
Date of Result Entered : **17/10/2022**
Date of Report Published : **03/11/2022**

Sl. No.	Parameter	Method	Manufacturer	1	2	3	4	5	6	7	8	9
1	GLUCOSE II	GOD-POD II	Transasia / Erba	854	363.74	10.40	37.83	387.4	0.63	2.59		
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	848	35.79	12.27	4.39	39	0.73	0.30		
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	555	1.49	14.04	0.21	1.3	-0.91	0.02		
4	BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	714	1.46	19.22	0.28	1.42	-0.14	0.02		
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	666	5.77	7.82	0.45	5.1	-1.49	0.03		
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	717	3.31	8.64	0.29	3.05	-0.91	0.02		
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	685	7.43	10.59	0.79	7.49	0.08	0.05		
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	790	115.96	10.02	11.62	101	-1.29	0.83		
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	827	210.76	13.41	28.26	230.2	0.69	1.97		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

SRI AMMAN MICRO LAB
OPP TO GH, 118, RS ROAD, PERUNDURAI
ERODE
TAMILNADU638052

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****