

VIEW LAB MONTHLY SUMMARY

Lab Name [SRI AMMAN MICRO LAB](#)[Details About Robust Analysis](#)Lab No [16124](#)[Detail About Monthly Summary](#)Month [September](#)[Detail about SDI](#)Year [2022](#)Constituent Group [Chemistry II](#)[Click on the analyte to view Graphical Data](#)[All Analyser Result](#)[Print](#)Date of Result Entered : [17/09/2022](#)Date of Report Published : [05/10/2022](#)

1	GLUCOSE II	GOD-POD II	Transasia / Erba	886	188.05	9.72	18.27	212 mg/dL	1.31	1.23
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	856	51.66	12.91	6.67	59.7 mg/dL	1.21	0.46
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	562	2.62	10.95	0.29	2.65 mg/dL	0.10	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	827	2.94	13.86	0.41	3.08 mg/dL	0.34	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	690	5.72	8.35	0.48	5.21 g/dL	-1.07	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	714	3.33	9.02	0.30	3.04 g/dL	-0.97	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	782	5.67	13.33	0.76	4.42 mg/dL	-1.65	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	789	109.84	10.64	11.68	107.6 mg/dL	-0.19	0.83
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	764	161.37	11.70	18.88	170 mg/dL	0.46	1.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Home	About Us	Contact Us	Privacy Policy
Sri Amman Micro Lab	Sri Amman Micro Lab	Sri Amman Micro Lab	Sri Amman Micro Lab
Sri Amman Micro Lab	Sri Amman Micro Lab	Sri Amman Micro Lab	Sri Amman Micro Lab