

VIEW LAB MONTHLY SUMMARY

Lab Name **SRI AMMAN MICRO LAB**Lab No **16124**Month **July**Year **2022**Constituent Group **Chemistry II**

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 20/07/2022

Date of Report Published : 05/08/2022

| Sl No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|-----------------|---|--|--------------------|--------|--------------|-------|-------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE II | GOD-POD II | Transasia / Erba | 841 | 313.52 | 11.25 | 35.28 | 245.9 mg/dL | -1.92 | 2.43 |
| 2 | UREA II | Others (Urease Berthelot / Ned Dye) II | Any Analyser (Automation / Semi Automation) | 863 | 58.70 | 12.57 | 7.38 | 137.4 mg/dL | 10.66 | 0.50 |
| 3 | CREATININE II | Jaffes Kinetic-Alkaline Picrate II | Transasia / Erba | 552 | 3.38 | 10.92 | 0.37 | 3.7 mg/dL | 0.87 | 0.03 |
| 4 | T.BILIRUBIN II | Diazonium Salt (Colorimetric) / Jendrassik II | Transasia / Erba | 756 | 1.82 | 18.13 | 0.33 | 2.44 mg/dL | 1.88 | 0.02 |
| 5 | T-PROTEIN II | Biuret - Colorimetric II | Transasia / Erba | 729 | 5.98 | 9.10 | 0.54 | 4.94 g/dL | -1.91 | 0.04 |
| 6 | ALBUMIN II | BCG - Colorimetric II | Transasia / Erba | 718 | 3.39 | 8.38 | 0.28 | 2.88 g/dL | -1.80 | 0.02 |
| 7 | URIC ACID II | Enzymatic / Uricase Colorimetric II | Transasia / Erba | 772 | 7.91 | 12.86 | 1.02 | 8.25 mg/dL | 0.33 | 0.07 |
| 8 | CHOLESTROL II | CHOD-PAP II | Transasia / Erba | 826 | 116.69 | 10.53 | 12.28 | 107.9 mg/dL | -0.72 | 0.85 |
| 9 | TRIGLYCERIDE II | GPO-PAP / Enzymatic Colorimetric / End Point II | Transasia / Erba | 750 | 253.80 | 11.48 | 29.13 | 238.7 mg/dL | -0.52 | 2.13 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |