

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001:2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,  
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



CSTIN: 33ADSPA9153Q1Z4 • DL No: TN-12-208-00398, TN-12-218-00398

## CALIBRATION CERTIFICATE

Customer Name & Address:	<b>M/S. Sri Amman Micro Lab,</b> #118, R.S. Road, Opp.Govt. Hopsital, Perundurai-638052	Certificate No:	AMS/LAB/289/2271
		Issuing Date:	12/09/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	<b>Micropipette</b>	Id. No:	<b>SAM/LAB/VMP/003</b>
Make:	Erba	Department:	Laboratory
Model:	---	Calibrated On:	<b>11/09/2022</b>
Sl. No:	OC499335	Calibration Due Suggested:	<b>10/09/2023</b>
Type:	Variable	DUC Condition on Receipt:	<b>OK</b>
Range & Least Count:	100 – 1000µl		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/22/ATL/0432-02	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

### Test Results Volume µl:

SL. No.	DUC Reading (µl)	Standard Reading (µl)		Correction (µl)
		Trial 1	Trial 2	
1	100.0	110.00	110.00	+10.00
2	200.0	210.00	210.00	+10.00
3	500.0	500.00	500.00	+00.00
4	800.0	800.00	800.00	+00.00
5	1000.0	990.00	990.00	-10.00

Allowable Tolerance: User Defined

### Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

**Calibrated By:**  
(Calibration Engineer)

aceagm@gmail.com



amshealthcaremduc@gmail.com

**Approved by:**  
(Technical Manager)

94434 74610, 93455 56148

