

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001:2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



CSTIN 33ADSPA9153Q1ZA • DI No TN-12 20B 00398, TN 12 21B 00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Sri Amman Micro Lab, #118, R.S. Road, Opp.Govt. Hopsital, Perundurai-638052	Certificate No:	AMS/LAB/289/2268
		Issuing Date:	12/09/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	SAM/LAB/CEN/001
Make:	Droplet	Department:	Laboratory
Model:	8-BKT	Calibrated On:	11/09/2022
Sl. No:	---	Calibration Due Suggested:	10/09/2023
Type:	Analog – Manual Control	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 3500 / LC – 1; Timer 10 to 60Mins. In 10Mins. interval		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

Test Results: Rotation Per Minute

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	Position 1	1420	0
2	Position 2	1800	0
3	Position 3	2120	0
4	Position 4	2440	0
5	Position 5	3010	0

Allowable Tolerance: Set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

aceagm@gmail.com



amshealthcaremdc@gmail.com

Approved by:

(Technical Manager)

94434 74610, 93455 56148