

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001:2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



UIN: 33ADSP69153012A • DL No: TN 12-20B-00398 TN 12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Sri Amman Micro Lab, #118, R.S. Road, Opp.Govt. Hopsital, Perundurai-638052	Certificate No:	AMS/LAB/289/2265
		Issuing Date:	12/09/2022
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	SAM/LAB/REFR/01
Make:	Whirlpool	Department:	Laboratory
Model:	Imagic-5g	Calibrated On:	11/09/2022
Sl. No:	1471330134105142	Calibration Due Suggested:	10/09/2023
Type:	Reagent	DUC Condition on Receipt:	OK
Range & Least Count:	-10 to 8°C/ LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	PC/01/22/ATL/0171-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023

Room Temperature: 25°C±2 Room Humidity: 30-60% (At the time of Calibration)

Test Results:

SL.No.	Place	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	Freezer	FRZ 1	-1.18	0
2		FRZ 2	-0.96	0
3	Chiller	CHI 1	1.13	0
4		CHI 2	1.08	0
5	Chamber I	CHA 1	3.47	0
6		CHA 2	3.51	0
7	Chamber II	CHA 1	4.33	0
8		CHA 2	4.28	0
9	Chamber III	CHA 1	5.13	0
10		CHA 2	5.24	0

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:
(Calibration Engineer)

aceagm@gmail.com



amshealthcaremd@gmail.com

Approved by:
(Technical Manager)

94434 74610, 93455 56148