

CALIBRATION CERTIFICATE

Ref No. TS/6618/SBP

Date : 5th sep'2022

**Arasu Diagnostics centre,
#1/4, Venkatrama Iyar Street,
Nadimuthu Nagar,
Pattukkottai-614602**

Dear Sir,

As desired by you please find the Calibration report

1. Year of manufacturing : 2017 Easy lyte Sr No. 54070CNKC
2. Date of Calibration - 2nd Sep'2022

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK/not OK
Logic 5 V	5.06V	OK
Line voltage	230V	OK
Logic 12V	12V	OK
Calibration Value (Na)	59.45	OK
Calibration Value (K)	59.71	OK

TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072
Tel.: (022) 40309000, Fax : (022) 2857 3030 Email :
transasia@transasia.co.in Website : www.transasia.co.in

CALIBRATION CERTIFICATE

Calibration Value (Li)	49.04	OK
Electrode Values(Na)	131	OK
Electrode Value(K)	85	OK
Electrode Value(Cl)	57	OK
Pump calibration		OK
Fluid Test		OK

.....2

Instrument checked with Randox (Bovine) Controls

Thanking you and assuring of our prompt attention at all times.



Yours faithfully,
For **TRANSASIA BIO-MEDICALS LTD.**

TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072
Tel.: (022) 40309000, Fax : (022) 2857 3030 Email :
transasia@transasia.co.in Website : www.transasia.co.in