



RML – Quality Assurance Program (RML – QAP)



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 2419

Cycle - 11/2022
Round No :10

Date: 20/10/2022

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	253	200	79%	1.010-1.020	1.020	Within Consensus
pH	-	260	211	81%	7.0-8.0	7.0	Within Consensus
Leukocytes	/µl	251	187	75%	2+·3+	2+	Within Consensus
Nitrite	-	249	249	86%	Positive	Positive	Within Consensus
Protein	mg/dl	258	189	73%	2+·3+	2+	Within Consensus
Glucose	mg/dl	256	224	88%	2+·4+	4+	Within Consensus
Ketones	-	258	224	87%	1+·3+	2+	Within Consensus
Urobilinogen	mg/dl	256	180	70%	2+·3+	2+	Within Consensus
Bilirubin	mg/dl	255	239	94%	2+·3+	3+	Within Consensus
Blood (Hemoglobin)	/µl	256	227	89%	3+·5+	3+	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#) Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By:

Prepared By:

Programme Director

Dr. Bandana Mehrotra

Page 1 of 4

Reviewed & found satisfactory
Dr. Mehrotra
5/11/2022

Doc No.: ASS / FR / 10 / R 01 / Dt: 05 01 2022



Address: B-171, Nirala Nagar, Lucknow - 226 020 Ph.: 403-4100-130 (30 Lines), 4077180, 2788444 Fax: (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence