



CALIBRATION CERTIFICATE

DATE	2/5/2022
CUSTOMER NAME	RIA DIAGNOSTIC CENTRE
ADDRESS	NAVARANG, RAJAJINAGAR, BANGALORE
INSTRUMENT	FULLY AUTOMATED BIOCHEMISTRY ANALYZER
SERIAL NUMBER	60843
MODEL	EM 360

VERIFICATION DETAILS

SL. NO.	OPTICAL FILTER	EXPECTED RANGE OF FILTER VALUES	MEASURED FILTER VALUES
1	340 nm	50-250	158
2	376 nm	50-250	175
3	415 nm	50-250	174
4	450 nm	50-250	156
5	480 nm	50-250	150
6	505 nm	50-250	177
7	546 nm	50-250	173
8	570 nm	50-250	176
9	600 nm	50-250	158
10	660 nm	50-250	178
11	700 nm	50-250	157
12	750 nm	50-250	159

Result were found to be in range

Control values for ERBA NORM Lot : S062125B EXP: 30-11-23

Calibrated Date : 2/5/22

Next Calibration Due on : 1/5/23

for TRANSASIA BIO-MEDICALS LTD.,

Technical Service Department

